

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Krogedal AGENCY PVPD
DATE 3/2/11 TIME 0645
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD Booking Room

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 095 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE CB Krogedal #90/QAS #33200

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

03/02/2011

06:50:34

Standard Lot#: 779386

Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90

PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90

PRESCOTT VALLEY PD

Test	g/210L	Time
Air Blank	0.000	06:51:14
0.100 Cal Check	0.095	06:51:35
Air Blank	0.000	06:52:05

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Krogedal AGENCY PVPD
DATE 3/2/11 TIME 0655
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD Booking

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 095 AC. Result 0. 094 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Cl Krogedal #90 / QAS # 33200

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number: 80-000293
Core Version: 8105.48
03/02/2011 06:53:23

Standard Lot#: 779386
Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST
DOB: 01/15/1975
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	06:54:25
Diagnostic Test	Pass	06:54:59
Air Blank	0.000	06:55:26
0.100 Cal Check	0.095	06:55:47
Air Blank	0.000	06:56:16
Subject Test	0.000	06:56:38
Air Blank	0.000	06:57:06
Wait	ABT*	
Air Blank	0.000	06:57:40

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number: 80-000293
Core Version: 8105.48
03/02/2011 06:57:46

Standard Lot#: 779386
Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST]
DOB: 01/15/1967
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	06:58:27
Air Blank	0.000	06:58:55

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

03/02/2011

06:59:00

Standard Lot#: 779386

Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90

PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90

PRESCOTT VALLEY PD

Subject: SQAP,TEST

DOB: 10/15/1957

Sex: M

Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	06:59:31
Air Blank	0.000	07:00:00

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT VALLEY PD
Serial Number: 80-000293
Core Version: 8105.48
03/02/2011 07:00:24

Standard Lot#: 779386
Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90
PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90
PRESCOTT VALLEY PD

Subject: SQAP,TEST
DOB: 10/15/1957
Sex: M Weight: 150

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:01:19
Diagnostic Test	Pass	07:01:53
Air Blank	0.000	07:02:20
0.100 Cal Check	0.094	07:02:41
Air Blank	0.000	07:03:11
Subject Test	0.150*	07:06:18
Air Blank	0.000	07:06:50
Air Blank	0.000	07:07:18
Subject Test	XXX**	07:07:37
Air Blank	0.000	07:08:15

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence