

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Krogedal AGENCY PVPD
DATE 6/13/11 TIME 2210
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD

- (C) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (C) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(C) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(C) 4. Air blank completed.
(C) 5. Calibration check completed. Test results 0. 096 AC.
(C) 6. Air blank completed.
(C) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE C. Krogedal #90 / QAS #33200

DPS Form Exh G-2 (Rev 05-01)

Cal. Check completed due to another officer reporting the instrument had ~~no~~ a blank screen and then displayed "Not Ready Mode" and would not go to "ready mode" to complete a test. This occurred on 6/12/11 and I completed this G-2 when I came back on duty. The instrument responded normally with no problems or abnormalities. I would surmise for an unknown reason the instrument went into "Not Ready Mode" which takes several minutes for the instrument to transition to "Ready Mode". The officer was not a QAS and was not familiar with this procedure.

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

06/13/2011

22:13:43

Standard Lot#: 779386

Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90

PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90

PRESCOTT VALLEY PD

Test	g/210L	Time
Air Blank	0.000	22:14:29
0.100 Cal Check	0.096	22:14:50
Air Blank	0.000	22:15:19

Preventative Maintenance
Performed -- Timer Reset