

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Krogedal AGENCY PVPO
DATE 11/23/11 TIME 1411
INTOXILYZER SERIAL # 80-000293 LOCATION PVPO

- (C) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution _____ AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (C) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (C) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (C) 4. Air blank completed.
- (C) 5. Calibration check completed. Test results 0. _____ AC.
- (C) 6. Air blank completed.
- (C) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Ces Krogedal #90 / QAS # 33200

DPS Form Exh G-2 (Rev 05-01)

G-2 performed after Standard Change

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

11/23/2011

14:11:59

Standard Lot#: 752434

Last Changed By: C. KROGEDAL #90

QAS: C. KROGEDAL #90

PRESCOTT VALLEY PD

Operator: C. KROGEDAL #90

PRESCOTT VALLEY PD

Test g/210L Time

Air Blank 0.000 14:12:30
0.100 Cal Check 0.094 14:12:51
Air Blank 0.000 14:13:21

Preventative Maintenance
Performed -- Timer Reset