

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. Lawrence AGENCY Prescott Valley PD  
DATE 12-23-15 TIME 2229  
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD / Booking Room

- ( 5 ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- ( 5 ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( 4 ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( 5 ) 4. Air blank completed.
- ( 5 ) 5. Calibration check completed. Test results 0. 098 AC.
- ( 4 ) 6. Air blank completed.
- ( 4 ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE \_\_\_\_\_

DPS Form Exh G-2 (Rev 05-01)

