

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. LAWRENCE AGENCY PRESOTT VALLEY PD
DATE 10-8-16 TIME 2239
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD / Booking Room

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 097 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

10/08/2016

22:39:24

Standard Lot#: 06616100A1

Last Changed By: A. LAWRENCE #61

QAS: A. LAWRENCE #61

PRESCOTT VALLEY PD

Operator: A. LAWRENCE #61

PRESCOTT VALLEY PD

Test	g/210L	Time
Air Blank	0.000	22:40:20
0.100 Cal Check	0.097	22:40:41
Air Blank	0.000	22:41:10

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST T. LAWRENCE AGENCY Prescott Valley PD
DATE 10-8-16 TIME 2241
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD / Booking Room

(41. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (41. Clock time check.
- (42. Date check.

OPERATIONAL TESTS

- (41. Alcohol-free subject test result 0. 000 AC.
- (42. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (43. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (44. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

Intox Working Correctly

SIGNATURE [Signature]

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

10/08/2016

22:41:54

Standard Lot#: 06616100A1

Last Changed By: A. LAWRENCE #61

QAS: A. LAWRENCE #61

PRESCOTT VALLEY PD

Operator: A. LAWRENCE #61

PRESCOTT VALLEY PD

Subject: SQAP, TEST

DOB: 01/05/1980

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:42:55
Diagnostic Test	Pass	22:43:29
Air Blank	0.000	22:43:56
0.100 Cal Check	0.097	22:44:17
Air Blank	0.000	22:44:46
Subject Test	0.000	22:45:10
Air Blank	0.000	22:45:37
Wait	ABT*	
Air Blank	0.000	22:46:12

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Firmware Version: 8105.48

10/08/2016

22:46:18

Standard Lot#: 06616100A1

Last Changed By: A. LAWRENCE #61

AS: A. LAWRENCE #61

PRESCOTT VALLEY PD

Operator: A. LAWRENCE #61

PRESCOTT VALLEY PD

Subject: SQAP, TEST

DOB: 01/05/1980

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	22:47:10
Air Blank	0.000	22:47:39

*Improper Sample



Not a Successfully Completed Test Sequence

INNOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

10/08/2016

22:47:45

Standard Lot#: 06616100A1

Last Changed By: A. LAWRENCE #61

CAS: A. LAWRENCE #61

PRESCOTT VALLEY PD

Operator: A. LAWRENCE #61

PRESCOTT VALLEY PD

Subject: SQAP, TEST

DOB: 01/05/1980

Len: N

Weight: 200

35 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	22:48:31
Air Blank	0.000	22:48:59

RFI Detect



Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

10/08/2016

22:49:02

Standard Lot#: 06616100A1

Last Changed By: A. LAWRENCE #61

QAS: A. LAWRENCE #61

PRESCOTT VALLEY PD

Operator: A. LAWRENCE #61

PRESCOTT VALLEY PD

Subject: SQAP, TEST

DOB: 01/05/1980

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:49:57
Diagnostic Test	Pass	22:50:31
Air Blank	0.000	22:50:59
0.100 Cal Check	0.097	22:51:20
Air Blank	0.000	22:51:49
Subject Test	0.136*	22:54:57
Air Blank	0.000	22:55:29
Air Blank	0.000	22:55:57
Subject Test	XXX**	22:56:23
Air Blank	0.000	22:56:57

*Insufficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. LAWRENCE AGENCY PRESOTT VALLEY PD
DATE 10-8-16 TIME 2257
INTOXILYZER SERIAL # 80000293 LOCATION PVPD 1 Booking Room

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 096 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

