

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST B. PENNER AGENCY PRESCOTT VALLEY PD  
DATE 8-16-17 TIME 1208  
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD MAIN-HOLDING

(~~1~~) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (~~1~~) 1. Clock time check.  
(~~2~~) 2. Date check.

OPERATIONAL TESTS

- (~~1~~) 1. Alcohol-free subject test result 0. 000 AC.  
(~~2~~) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(~~3~~) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(~~4~~) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

---

---

---

---

SIGNATURE  #29

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

08/16/2017

12:08:35

---

Standard Lot#: 06117100A2

Last Changed By: B. PENNER #29

---

QAS: B. PENNER #29

PRESCOTT VALLEY PD

---

Operator: B. PENNER #29

PRESCOTT VALLEY PD

---

Subject: SQAP, TEST

DOB: 01/01/2017

Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:10:41
Diagnostic Test	Pass	12:11:15
Air Blank	0.000	12:11:42
0.100 Cal Check	0.098	12:12:03
Air Blank	0.000	12:12:33
Subject Test	0.000	12:12:51
Air Blank	0.000	12:13:19
Wait	ABT*	
Air Blank	0.000	12:13:57

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

08/16/2017

12:15:19

---

Standard Lot#: 06117100A2

Last Changed By: B. PENNER #29

---

QAS: B. PENNER #29

PRESCOTT VALLEY PD

---

Operator: B. PENNER #29

PRESCOTT VALLEY PD

---

Subject: SQAP,TEST

DOB: 01/01/2017

Sex: M

Weight: 112

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	12:16:31
Air Blank	RFI**	12:16:49

\*Improper Sample

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

08/16/2017

12:18:38

---

Standard Lot#: 06117100A2

Last Changed By: B. PENNER #29

---

QAS: B. PENNER #29

PRESCOTT VALLEY PD

---

Operator: B. PENNER #29

PRESCOTT VALLEY PD

---

Subject: SQAP,TEST

DOB: 01/01/2017

Sex: M

Weight: 113

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:19:49
Diagnostic Test	Pass	12:20:23
Air Blank	0.000	12:20:51
0.100 Cal Check	0.097	12:21:11
Air Blank	0.000	12:21:41
Subject Test	0.000*	12:24:48
Air Blank	0.000	12:25:21
Air Blank	0.000	12:25:49
Subject Test	XXX**	12:26:30
Air Blank	0.000	12:27:03

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST B. PENNER AGENCY PRESOTT VALLEY PD  
DATE 8-16-17 TIME 1230  
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD MAIN-HOLDING

- (H) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
~~OR~~  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (H) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (H) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (H) 4. Air blank completed.
- (H) 5. Calibration check completed. Test results 0. 097 AC.
- (H) 6. Air blank completed.
- (H) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

POST SQAP

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

08/16/2017

12:30:20

-----  
Standard Lot#: 06117100A2

Last Changed By: B. PENNER #29  
-----

QAS: B. PENNER #29

PRESCOTT VALLEY PD  
-----

Operator: B. PENNER #29

PRESCOTT VALLEY PD  
-----

Test	g/210L	Time
Air Blank	0.000	12:30:51
0.100 Cal Check	0.097	12:31:11
Air Blank	0.000	12:31:41

Preventative Maintenance  
Performed -- Timer Reset