

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST B. PENNER #29 AGENCY PRESOTT VALLEY PD
DATE 10-13-17 TIME 0946
INTOXILYZER SERIAL # 80-000293 LOCATION PVPA MAIN - HOLDING

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .10 AC.
~~OR~~
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator.
Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34 ± 0.2 C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  #29

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

10/13/2017

09:49:33

Standard Lot#: 06117100A2

Last Changed By: B. PENNER #29

QAS: B. PENNER #29

PRESCOTT VALLEY PD

Operator: B. PENNER #29

PRESCOTT VALLEY PD

Test	g/210L	Time
Air Blank	0.000	09:50:04
0.100 Cal Check	0.099	09:50:25
Air Blank	0.000	09:50:55

Preventative Maintenance
Performed -- Timer Reset