

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST B. PENNER AGENCY PRESOTT VALLEY PD
DATE 12-6-18 TIME 0952
INTOXILYZER SERIAL # 80-000293 LOCATION AVPD MAIN - HOLDING

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator.

Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34 ± 0.2 C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  #29

DPS Form Exh G-2 (Rev 05-01)

PRE-SQAP

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

12/06/2018

09:55:15

Standard Lot#: 10818100A2

Last Changed By: T. TEST #0000

QAS: B. PENNER #29

PRESCOTT VALLEY PD

Operator: B. PENNER #29

PRESCOTT VALLEY PD

Test	g/210L	Time
Air Blank	0.000	09:56:06
0.100 Cal Check	0.099	09:56:27
Air Blank	0.000	09:56:56

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST B. PENNER AGENCY PRESOTT VALLEY PD
DATE 12-6-18 TIME 1001
INTOXILYZER SERIAL # 80-000293 LOCATION PVPD MAIN-HOLDING

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE  #29

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

12/06/2018

10:01:12

Standard Lot#: 10818100A2

Last Changed By: T. TEST

#0000

QAS: B. PENNER

#29

PRESCOTT VALLEY PD

Operator: B. PENNER

#29

PRESCOTT VALLEY PD

Subject: SQAP, TEST

DOB: 01/01/2018

Sex: M

Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:02:53
Diagnostic Test	Pass	10:03:27
Air Blank	0.000	10:03:55
0.100 Cal Check	0.098	10:04:16
Air Blank	0.000	10:04:45
Subject Test	0.000	10:05:04
Air Blank	0.000	10:05:32
Wait	ABT*	
Air Blank	0.000	10:06:10

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

12/06/2018

10:06:35

Standard Lot#: 10818100A2

Last Changed By: T. TEST

#0000

QAS: B. PENNER

#29

PRESCOTT VALLEY PD

Operator: B. PENNER

#29

PRESCOTT VALLEY PD

Subject: SQAP, TEST

DOB: 01/01/2018

Sex: M

Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	10:07:36
Air Blank	0.000	10:08:04

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

12/06/2018

10:08:35

Standard Lot#: 10818100A2

Last Changed By: T. TEST

#0000

QAS: B. PENNER

#29

PRESCOTT VALLEY PD

Operator: B. PENNER

#29

PRESCOTT VALLEY PD

Subject: SQAP,TEST

DOB: 01/01/2018

Sex: M

Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:10:05
Air Blank	0.000	10:10:34

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

12/06/2018

10:11:06

Standard Lot#: 10818100A2

Last Changed By: T. TEST

#0000

QAS: B. PENNER

#29

PRESCOTT VALLEY PD

Operator: B. PENNER

#29

PRESCOTT VALLEY PD

Subject: SQAP,TEST

DOB: 10/10/2018

Sex: M

Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:12:39
Diagnostic Test	Pass	10:13:12
Air Blank	0.000	10:13:40
0.100 Cal Check	0.098	10:14:01
Air Blank	0.000	10:14:30
Subject Test	0.009*	10:17:37
Air Blank	0.000	10:18:10
Air Blank	0.000	10:18:37
Subject Test	XXX**	10:18:53
Air Blank	0.000	10:19:38

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST B. PENNER AGENCY PRESOTT VALLEY PD
DATE 12-6-18 TIME 1020
INTOXILYZER SERIAL # 80-000293 LOCATION PVPO MAIN-HOLDINGS

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
~~OR~~
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator.
Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34 \cdot C \pm 0.2 \cdot C$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(✓) 4. Air blank completed.
(✓) 5. Calibration check completed. Test results 0. 098 AC.
(✓) 6. Air blank completed.
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  #26

DPS Form Exh G-2 (Rev 05-01)

POST SEAL

INTOXILYZER 8000

Location: PRESCOTT VALLEY PD

Serial Number: 80-000293

Core Version: 8105.48

12/06/2018

10:21:56

Standard Lot#: 10818100A2

Last Changed By: T. TEST #0000

QAS: B. PENNER #29

PRESCOTT VALLEY PD

Operator: B. PENNER #29

PRESCOTT VALLEY PD

Test	g/210L	Time
Air Blank	0.000	10:22:34
0.100 Cal Check	0.098	10:22:54
Air Blank	0.000	10:23:24

Preventative Maintenance
Performed -- Timer Reset