


EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST G. KOESTER AGENCY PRESOTT PD  
DATE 5-8-08 TIME 1104  
INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT PD

- (GK) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (GK) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (GK) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (GK) 4. Air blank completed.
- (GK) 5. Calibration check completed. Test results 0. 099 AC.
- (GK) 6. Air blank completed.
- (GK) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE G. KOESTER  12/2

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

05/08/2008

11:03:25

-----  
Standard Lot#: 727501I

Last Changed By: T. ATTENBERGER #252

-----  
QAS: G. KOESTER #212

PRESCOTT PD  
-----

Test	g/210L	Time
Air Blank	0.000	11:03:52
Cal Check	0.099	11:04:11
Air Blank	0.000	11:04:41

G. KOESTER 

Operator's Signature

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST G. KOESTER AGENCY PRESCOTT PD

DATE 5-8-08 TIME 1105

INTOXILYZER SERIAL # 80-000337 LOCATION PRESCOTT PD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.

2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 0.158 AC.

2. Error recognition logic system functioning.

Not a Successfully Completed Test Sequence printed

3. Proper sample recognition system.

Not a Successfully Completed Test Sequence printed

Deficient sample printed.

4. Standard Calibration Check standard 0. 0.100 AC. Result 0. 0.098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

ALL SYSTEMS TESTS, NO ERRORS

SIGNATURE G. KOESTER  5-20

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
05/08/2008

11:06:08

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: G. KOESTER #212  
PRESCOTT PD

---

Operator: G. KOESTER #212  
PRESCOTT PD

---

Subject: SQAP,TEST  
DOB: 01/01/1901  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	11:07:08
Diagnostic Test	Pass	11:07:35
Air Blank	0.000	11:08:03
0.100 Cal Check	0.098	11:08:21
Air Blank	0.000	11:08:51
Subject Test	0.000	11:09:11
Air Blank	0.000	11:09:38
Five Minute Wait	ABT*	
Air Blank	0.000	11:10:16

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
05/08/2008

11:10:49

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: G. KOESTER #212  
PRESCOTT PD

---

Operator: G. KOESTER #212  
PRESCOTT PD

---

Subject: SQAP,TEST  
DOB: 01/01/1901  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	11:11:42
Air Blank	0.000	11:12:11

\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
05/08/2008

11:12:25

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: G. KOESTER #212  
PRESCOTT PD

---

Operator: G. KOESTER #212  
PRESCOTT PD

---

Subject: SQAP,TEST  
DOB: 01/01/1901  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	11:12:57
Air Blank	0.000	11:13:25

\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
05/08/2008

11:13:58

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: G. KOESTER #212  
PRESCOTT PD

---

Operator: G. KOESTER #212  
PRESCOTT PD

---

Subject: SQAP, TEST  
DOB: 01/01/1901  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	11:15:00
Diagnostic Test	Pass	11:15:27
Air Blank	0.000	11:15:54
0.100 Cal Check	0.098	11:16:13
Air Blank	0.000	11:16:43
Subject Test	0.066*	11:19:51
Air Blank	0.000	11:20:24
Air Blank	0.000	11:20:52
Subject Test	XXX**	11:21:17
Air Blank	0.000	11:21:50

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST G. KOESTER AGENCY PRESCOTT PD  
DATE 5-8-08 TIME 1123  
INTOXILYZER SERIAL # 20-000337 LOCATION PRESCOTT PD

- (GK) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (GK) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (GK) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (GK) 4. Air blank completed.
- (GK) 5. Calibration check completed. Test results 0.098 AC.
- (GK) 6. Air blank completed.
- (GK) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE G. KOESTER De 10 . 212

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

05/08/2008

11:23:05

-----  
Standard Lot#: 727501I

Last Changed By: T. ATTENBERGER #252

-----  
QAS: G. KOESTER #212

PRESCOTT PD  
-----

Test	g/210L	Time
Air Blank	0.000	11:23:33
Cal Check	0.098	11:23:52
Air Blank	0.000	11:24:21

G. KOESTER 

Operator's Signature