

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. Arsenberger #252 AGENCY Prescott P.D.  
DATE 073108 TIME 1246  
INTOXILYZER SERIAL # 80-000337 LOCATION P.P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
() 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
() 4. Air blank completed.  
() 5. Calibration check completed. Test results 0. 097 AC.  
() 6. Air blank completed.  
() 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE T. Arsenberger #252

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

07/31/2008

12:46:33

-----  
Standard Lot#: 727501I

Last Changed By: T. ATTENBERGER #252  
-----

QAS: T. ATTENBERGER #252

PRESCOTT PD  
-----

Test	g/210L	Time
Air Blank	0.000	12:47:01
Cal Check	0.097	12:47:19
Air Blank	0.000	12:47:49

  
-----  
Operator's Signature

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST T. ATMANBERGER #20 AGENCY PRESCOTT P.D.  
DATE 073108 TIME 1249  
INTOXILYZER SERIAL # 80-000337 LOCATION P.P.D.

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.  
(✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.  
(✓) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(✓) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

I BLEW TOO LONG DURING PROPER SAMPLE RECOG / "DEFICIENT  
SAMPLE" AND SAMPLE TOOK RE-DID  
\* ALL FUNCTIONING PROPERLY.

SIGNATURE T. Atmanberger #20

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
07/31/2008

12:49:51

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: T. ATTENBERGER #252  
PRESCOTT PD

---

Operator: T. ATTENBERGER #252  
PRESCOTT PD

---

Subject: SQAP, TEST  
DOB: 01/01/1900  
Sex: M Weight: 300

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:51:19
Diagnostic Test	Pass	12:51:47
Air Blank	0.000	12:52:14
0.100 Cal Check	0.097	12:52:33
Air Blank	0.000	12:53:02
Subject Test	0.000	12:53:35
Air Blank	0.000	12:54:02
Five Minute Wait	ABT*	
Air Blank	0.000	12:55:10

\*Sequence Aborted

Handwritten signature and initials, possibly "T. Attenberger #252".

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
07/31/2008

12:56:37

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: T. ATTENBERGER #252  
PRESCOTT PD

---

Operator: T. ATTENBERGER #252  
PRESCOTT PD

---

Subject: SQAP, TEST  
DOB: 01/01/1900  
Sex: M Weight: 300


---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	12:57:23
Air Blank	0.000	12:57:51

\*Improper Sample



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
07/31/2008 12:59:01

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: T. ATTENBERGER #252  
PRESCOTT PD

---

Operator: T. ATTENBERGER #252  
PRESCOTT PD

---

Subject: SQAP, TEST  
DOB: 01/01/1900  
Sex: M Weight: 300

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	12:59:54
Air Blank	0.000	13:00:22

\*RFI Detect



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
07/31/2008 13:01:02

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: T. ATTENBERGER #252  
PRESCOTT PD

---

Operator: T. ATTENBERGER #252  
PRESCOTT PD

---

Subject: SQAP, TEST  
DOB: 01/01/1900  
Sex: M Weight: 300

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:02:21
Diagnostic Test	Pass	13:02:48
Air Blank	0.000	13:03:15
0.100 Cal Check	0.096	13:03:34
Air Blank	0.000	13:04:04
Subject Test	0.000	13:04:31
Air Blank	0.000	13:04:59
Five Minute Wait		
Air Blank	0.000	13:09:59
Subject Test	0.000*	13:13:07
Air Blank	0.000	13:13:39

\*Deficient Sample

*Blow too long & it took.*



**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
07/31/2008

13:14:30

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: T. ATTENBERGER #252  
PRESCOTT PD

---

Operator: T. ATTENBERGER #252  
PRESCOTT PD

---

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 300


---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:15:39
Diagnostic Test	Pass	13:16:06
Air Blank	0.000	13:16:33
0.100 Cal Check	0.097	13:16:52
Air Blank	0.000	13:17:22
Subject Test	XXX*	13:17:50
Air Blank	0.000	13:18:24

\*Invalid Sample  
Begin new deprivation period



**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. AMEN BAZGAR #252 AGENCY PRESCOTT P.D.

DATE 073108 TIME 1300

INTOXILYZER SERIAL # 80-000337 LOCATION P.P.D.

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select (D) for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 096 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE T. Amen Bazgar #252

PRESCOTT P.D CHANGE 5.12.06

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

07/31/2008

13:20:36


-----  
Standard Lot#: 727501I

Last Changed By: T. ATTENBERGER #252  
-----

QAS: T. ATTENBERGER #252

PRESCOTT PD  
-----

Test	g/210L	Time
Air Blank	0.000	13:21:04
Cal Check	0.096	13:21:22
Air Blank	0.000	13:21:52

  
-----  
Operator's Signature