


EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST G. ROESTER AGENCY PRESOTT PD  
DATE 8-28-08 TIME 1533  
INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT

- (G) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (G) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (G) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (G) 4. Air blank completed.
- (G) 5. Calibration check completed. Test results 0. .096 AC.
- (G) 6. Air blank completed.
- (G) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  G. ROESTER

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

08/28/2008

15:36:14


-----  
Standard Lot#: 727501I

Last Changed By: T. ATTENBERGER #252

-----  
QAS: G. KOESTER #212

PRESCOTT PD  
-----

Test	g/210L	Time
Air Blank	0.000	15:36:41
Cal Check	0.096	15:37:00
Air Blank	0.000	15:37:30



Operator's Signature

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST G. KOESTER AGENCY PRESCOTT PD

DATE 8-28-08 TIME 1540

INTOXILYZER SERIAL # 80-050337 LOCATION PRESCOTT PD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.

2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 0.00 AC.

2. Error recognition logic system functioning.

Not a Successfully Completed Test Sequence printed

3. Proper sample recognition system.

Not a Successfully Completed Test Sequence printed

Deficient sample printed.

4. Standard Calibration Check standard 0. +150 AC. Result 0. -095 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

ALL SYSTEMS FUNCTIONING PROPERLY.

SIGNATURE G. KOESTER

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
08/28/2008

15:43:06

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: G. KOESTER #212  
PRESCOTT PD

---

Operator: G. KOESTER #212  
PRESCOTT PD

---

Subject: SQAP, TEST  
DOB: 01/01/1901  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	15:43:54
Diagnostic Test	Pass	15:44:21
Air Blank	0.000	15:44:49
0.100 Cal Check	0.096	15:45:08
Air Blank	0.000	15:45:37
Subject Test	0.000	15:45:54
Air Blank	0.000	15:46:22
Five Minute Wait	ABT*	
Air Blank	0.000	15:46:58

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
08/28/2008 15:47:39

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: G. KOESTER #212  
PRESCOTT PD

---

Operator: G. KOESTER #212  
PRESCOTT PD

---

Subject: SQAP,TEST  
DOB: 01/01/1901  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	15:48:11
Air Blank	0.000	15:48:39

\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
08/28/2008

15:48:44

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: G. KOESTER #212  
PRESCOTT PD

---

Operator: G. KOESTER #212  
PRESCOTT PD

---

Subject: SQAP,TEST  
DOB: 01/01/1901  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	15:49:18
Air Blank	0.000	15:49:46

\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
08/28/2008 15:50:12

---

Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

---

QAS: G. KOESTER #212  
PRESCOTT PD

---

Operator: G. KOESTER #212  
PRESCOTT PD

---

Subject: SQAP,TEST  
DOB: 01/01/1901  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	15:51:03
Diagnostic Test	Pass	15:51:30
Air Blank	0.000	15:51:58
0.100 Cal Check	0.097	15:52:17
Air Blank	0.000	15:52:46
Subject Test	0.058*	15:55:55
Air Blank	0.000	15:56:26
Air Blank	0.000	15:56:53
Subject Test	XXX**	15:57:26
Air Blank	0.000	15:57:58

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST G. KOESTER AGENCY ARIZONA SCOTT PD

DATE 8-22-08 TIME 1654

INTOXILYZER SERIAL # 80-000337 LOCATION ARIZONA SCOTT PD

- (GK) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (GK) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (GK) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (GK) 4. Air blank completed.
- (GK) 5. Calibration check completed. Test results 0. .095 AC.
- (GK) 6. Air blank completed.
- (GK) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  G. KOESTER +22

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number: 80-000337  
08/28/2008 16:04:08

-----  
Standard Lot#: 727501I  
Last Changed By: T. ATTENBERGER #252

-----  
QAS: G. KOESTER #212  
PRESCOTT PD

-----  
Test g/210L Time  
-----  
Air Blank 0.000 16:04:35  
Cal Check 0.095 16:04:54  
Air Blank 0.000 16:05:23

G. KOESTER 

Operator's Signature