

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. MONTGOMERY AGENCY Prescott P.D.  
DATE 1/20/08 TIME 0900  
INTOXILYZER SERIAL # 80-000337 LOCATION P.P.D.

- ( ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution 100 AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- ( ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
( ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
( ) 4. Air blank completed.  
( ) 5. Calibration check completed. Test results 0. 100 AC.  
( ) 6. Air blank completed.  
( ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE T. Montgomery

Prescott PD CHANGE 5-18-06

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number: 80-000337  
11/20/2008 09:00:08

-----  
Standard Lot#: 801002I  
Last Changed By: T. ATTENBERGER #252  
-----

QAS: T. ATTENBERGER #252  
PRESCOTT PD  
-----

Test	g/210L	Time
Air Blank	0.000	09:00:35
Cal Check	0.100	09:00:54
Air Blank	0.000	09:01:23



-----  
Operator's Signature

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST T. A. WAGNER #252 AGENCY Prescott P.D.

DATE 112008 TIME 0907

INTOXILYZER SERIAL # 80-000337 LOCATION P.P.D.

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.  
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0, 000 AC.  
() 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
() 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
() 4. Standard Calibration Check standard 0, 100 AC. Result 0, 100 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

ALL FUNCTIONING CORRECTLY.

SIGNATURE T. A. Wagner #252

DPS Form Exh G-4 (Rev 05-01)

PRESCOTT PD CHANGE 5 18 06

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
11/20/2008 09:07:13

Standard Lot#: 801002I  
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252  
PRESCOTT PD

Operator: T. ATTENBERGER #252  
PRESCOTT PD

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:08:26
Diagnostic Test	Pass	09:08:53
Air Blank	0.000	09:09:20
0.100 Cal Check	0.100	09:09:39
Air Blank	0.000	09:10:09
Subject Test	0.000	09:10:31
Air Blank	0.000	09:10:59
Five Minute Wait	ABT*	
Air Blank	0.000	09:11:37

\*Sequence Aborted



Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
11/20/2008

09:12:14

Standard Lot#: 801002I  
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252  
PRESCOTT PD

Operator: T. ATTENBERGER #252  
PRESCOTT PD

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	09:13:00
Air Blank	0.000	09:13:29

\*Improper Sample



Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number : 80-000337

11/20/2008

09:13:57

Standard Lot#: 801002I

Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252

PRESCOTT PD

Operator: T. ATTENBERGER #252

PRESCOTT PD

Subject: SQAP, TEST

DOB: 01/01/1900

Sex: M

Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	09:14:48
Air Blank	0.000	09:15:16

\*RFI Detect

Handwritten signature and initials, possibly "T. Attenberger" and "4-20-08".

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: PRESCOTT P.D  
Serial Number : 80-000337  
11/20/2008 09:16:06

Standard Lot#: 801002I  
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252  
PRESCOTT PD

Operator: T. ATTENBERGER #252  
PRESCOTT PD

Subject: SQAP,TEST  
DOB: 01/01/1900  
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:17:02
Diagnostic Test	Pass	09:17:30
Air Blank	0.000	09:17:57
0.100 Cal Check	0.100	09:18:16
Air Blank	0.000	09:18:45
Subject Test	0.015*	09:21:54
Air Blank	0.000	09:22:23
Air Blank	0.000	09:22:51
Subject Test	XXX**	09:23:24
Air Blank	0.000	09:23:56

\*Deficient Sample  
\*\*Invalid Sample  
Begin new deprivation period



Not a Successfully Completed Test Sequence

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. ATTENBOROUGH #252 AGENCY Prescott P.D.  
DATE 1/2008 TIME 0926  
INTOXILYZER SERIAL # 80-000337 LOCATION P.P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE T. Attenuborough #252

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

11/20/2008

09:26:13


-----  
Standard Lot#: 801002I

Last Changed By: T. ATTENBERGER #252  
-----

QAS: T. ATTENBERGER #252

PRESCOTT PD  
-----

Test	g/210L	Time
Air Blank	0.000	09:26:40
Cal Check	0.100	09:26:59
Air Blank	0.000	09:27:28



Operator's Signature