

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)


ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST G. KOESTER AGENCY PRESOTT PD

DATE 12-17-09 TIME 0747

INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT PD

- (K) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (K) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (K) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (K) 4. Air blank completed.
- (K) 5. Calibration check completed. Test results 0. .099 AC.
- (K) 6. Air blank completed.
- (K) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE G. KOESTER 

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

12/17/2009

07:47:03

Standard Lot#: 818201I

Last Changed By: G. KOESTER #212

QAS: G. KOESTER #212

PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	07:47:30
Cal Check	0.099	07:47:49
Air Blank	0.000	07:48:19

G. KOESTER 
Operator's Signature

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST G. KOESTER AGENCY PRESOTT PD
DATE 12-17-09 TIME 0750
INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT PD

~~(X)~~ 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ~~(X)~~ 1. Clock time check.
~~(X)~~ 2. Date check.

OPERATIONAL TESTS

- ~~(X)~~ 1. Alcohol-free subject test result 0. .000 AC.
~~(X)~~ 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
~~(X)~~ 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
~~(X)~~ 4. Standard Calibration Check standard 0. .100 AC. Result 0. .098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

ALL SYSTEMS FUNCTIONING PROPERLY

SIGNATURE G. KOESTER DKK

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number : 80-000337
12/17/2009

07:49:17

Standard Lot#: 818201I
Last Changed By: G. KOESTER #212

QAS: G. KOESTER #212
PRESCOTT PD

Operator: G. KOESTER #212
PRESCOTT PD

Subject: SQAP,TEST
DOB: 01/01/1901
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:50:10
Diagnostic Test	Pass	07:50:37
Air Blank	0.000	07:51:04
0.100 Cal Check	0.100	07:51:23
Air Blank	0.000	07:51:52
Subject Test	0.000	07:52:19
Air Blank	0.000	07:52:46
Five Minute Wait	ABT*	
Air Blank	0.000	07:53:23

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number : 80-000337
12/17/2009

07:53:43

Standard Lot#: 818201I
Last Changed By: G. KOESTER #212

QAS: G. KOESTER #212
PRESCOTT PD

Operator: G. KOESTER #212
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1901
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	07:54:11
Air Blank	0.000	07:54:40

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number : 80-000337
12/17/2009

07:54:56

Standard Lot#: 818201I
Last Changed By: G. KOESTER #212

QAS: G. KOESTER #212
PRESCOTT PD

Operator: G. KOESTER #212
PRESCOTT PD

Subject: SQAP,TEST
DOB: 01/01/1901
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	07:55:24
Air Blank	0.000	07:55:52

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number : 80-000337
12/17/2009

07:56:07

Standard Lot#: 818201I
Last Changed By: G. KOESTER #212

QAS: G. KOESTER #212
PRESCOTT PD

Operator: G. KOESTER #212
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1901
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:57:00
Diagnostic Test	Pass	07:57:27
Air Blank	0.000	07:57:55
0.100 Cal Check	0.099	07:58:14
Air Blank	0.000	07:58:43
Subject Test	0.042*	08:01:51
Air Blank	0.000	08:02:23
Air Blank	0.000	08:02:50
Subject Test	XXX**	08:03:41
Air Blank	0.000	08:04:17

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)


ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST G. KOESTER AGENCY PRESOTT PD

DATE 12-17-09 TIME _____

INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT PD

- (GK) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (GK) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (GK) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (GK) 5. Calibration check completed. Test results 0. 0.098 AC.
- (GK) 6. Air blank completed.
- (GK) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE G. KOESTER 

Prescott PD CHANGE 5-12-06

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

12/17/2009

08:07:05

Standard Lot#: 818201I

Last Changed By: G. KOESTER #212

QAS: G. KOESTER #212

PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	08:07:32
Cal Check	0.098	08:07:51
Air Blank	0.000	08:08:20

G. KOESTER 

Operator's Signature