

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST G. KOESTER AGENCY PRESTON PD
DATE 3-10-11 TIME 0925
INTOXILYZER SERIAL # 80-000337 LOCATION PRESTON PD

- (G) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (G) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (G) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (G) 4. Air blank completed.
- (G) 5. Calibration check completed. Test results 0.100 AC.
- (G) 6. Air blank completed.
- (G) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE G. KOESTER

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

03/10/2011

09:25:06

Standard Lot#: 25310100A1

Last Changed By: T. ATTENBERGER #252

QAS: G. KOESTER #212

PRESCOTT PD

Operator: G. KOESTER #212

PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	09:25:37
0.100 Cal Check	0.100	09:25:57
Air Blank	0.000	09:26:27

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST G. KOESTER AGENCY PRESOTT PD

DATE 3-10-11 TIME 0928

INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT PD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.

2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 0.000 AC.

2. Error recognition logic system functioning.

Not a Successfully Completed Test Sequence printed

3. Proper sample recognition system.

Not a Successfully Completed Test Sequence printed

Deficient sample printed.

4. Standard Calibration Check standard 0. 0.100 AC. Result 0. 0.096 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

ALL SYSTEMS FUNCTIONING PROPERLY

SIGNATURE G. KOESTER 

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

03/10/2011

09:27:58

Standard Lot#: 25310100A1

Last Changed By: T. ATTENBERGER #252

QAS: G. KOESTER #212
PRESCOTT PD

Operator: G. KOESTER #212
PRESCOTT PD

Subject: SQAP, TEST

DOB: 01/01/1980

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:29:01
Diagnostic Test	Pass	09:29:35
Air Blank	0.000	09:30:02
0.100 Cal Check	0.099	09:30:23
Air Blank	0.000	09:30:52
Subject Test	0.000	09:31:10
Air Blank	0.000	09:31:37
Wait	ABT*	
Air Blank	0.000	09:32:12

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
03/10/2011 09:32:17

Standard Lot#: 25310100A1
Last Changed By: T. ATTENBERGER #252

QAS: G. KOESTER #212
PRESCOTT PD

Operator: G. KOESTER #212
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1980
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	09:33:00
Air Blank	0.000	09:33:29

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
03/10/2011 09:33:49

Standard Lot#: 25310100A1
Last Changed By: T. ATTENBERGER #252

QAS: G. KOESTER #212
PRESCOTT PD

Operator: G. KOESTER #212
PRESCOTT PD

Subject: SQAP,TEST
DOB: 01/01/1980
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	09:34:20
Air Blank	0.000	09:34:48

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

03/10/2011

09:34:53

Standard Lot#: 25310100A1

Last Changed By: T. ATTENBERGER #252

QAS: G. KOESTER #212

PRESCOTT PD

Operator: G. KOESTER #212

PRESCOTT PD

Subject: SQAP,TEST

DOB: 01/01/1980

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:35:39
Diagnostic Test	Pass	09:36:13
Air Blank	0.000	09:36:41
0.100 Cal Check	0.099	09:37:02
Air Blank	0.000	09:37:31
Subject Test	0.068*	09:40:38
Air Blank	0.000	09:41:10
Air Blank	0.000	09:41:38
Subject Test	XXX**	09:41:57
Air Blank	0.000	09:42:31

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST G. KOESTER AGENCY PRESOTT PD

DATE 3-16-11 TIME 0930

INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT PD

- (G) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (G) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (G) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (G) 4. Air blank completed.
- (G) 5. Calibration check completed. Test results 0. .096 AC.
- (G) 6. Air blank completed.
- (G) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE G. KOESTER 

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

03/10/2011

09:42:58

Standard Lot#: 25310100A1

Last Changed By: T. ATTENBERGER #252

QAS: G. KOESTER #212

PRESCOTT PD

Operator: G. KOESTER #212

PRESCOTT PD

Test g/210L Time

Air Blank 0.000 09:43:28
0.100 Cal Check 0.096 09:43:49
Air Blank 0.000 09:44:18

Preventative Maintenance
Performed -- Timer Reset