

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. AIGENBERGER #252 AGENCY PHOENIX P.D.

DATE 10/21/11 TIME 1351

INTOXILYZER SERIAL # 80-000337 LOCATION PPD.

- (5) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. .100 AC.
OR
~~Pour a standard alcohol concentration solution into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C.~~
- (5) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (5) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (5) 4. Air blank completed.
- (5) 5. Calibration check completed. Test results 0. .100 AC.
- (5) 6. Air blank completed.
- (5) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
10/21/2011 13:51:34

Standard Lot#: 01011100A1
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	13:52:05
0.100 Cal Check	0.100	13:52:26
Air Blank	0.000	13:52:55

Preventative Maintenance
Performed -- Timer Reset



EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST T. Anderson AGENCY Peoria P.D.
DATE 10-21-11 TIME 1354
INTOXILYZER SERIAL # 80-000337 LOCATION PPD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 060 AC.
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 039 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

ALL FUNCTIONS WORK CORRECTLY.

SIGNATURE [Signature]

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
10/21/2011

13:54:17

Standard Lot#: 01011100A1
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1900 Weight: 300
Sex: M

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:55:26
Diagnostic Test	Pass	13:56:00
Air Blank	0.000	13:56:28
0.100 Cal Check	0.100	13:56:49
Air Blank	0.000	13:57:18
Subject Test	0.000	13:57:50
Air Blank	0.000	13:58:18
Wait	ABT*	
Air Blank	0.000	13:58:54

*Sequence Aborted



Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

10/21/2011

13:59:32

Standard Lot#: 01011100A1

Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP, TEST

DOB: 01/01/1900

Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	14:00:17
Air Blank	0.000	14:00:45

* Improper Sample



Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

10/21/2011

14:01:09

Standard Lot#: 01011100A1

Last Changed By: T. ATTENBERGER

#252

QAS: T. ATTENBERGER
PRESCOTT PD

#252

Operator: T. ATTENBERGER
PRESCOTT PD

#252

Subject: SQAP,TEST

DOB: 01/01/1900


Sex: M

Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	14:01:53
Air Blank	0.000	14:02:22

*RFI Detect



Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
10/21/2011

14:03:01

Standard Lot#: 01011100A1
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1900
Sex: M Weight: 300


15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:03:53
Diagnostic Test	Pass	14:04:27
Air Blank	0.000	14:04:54
0.100 Cal Check	0.099	14:05:15
Air Blank	0.000	14:05:45
Subject Test	0.073*	14:08:52
Air Blank	0.000	14:09:24
Air Blank	0.000	14:09:51
Subject Test	XXX**	14:10:21
Air Blank	0.000	14:10:54

*Deficient Sample

**Invalid Sample

Begin new deprivation period



Not a Successfully Completed Test Sequence

INTOXIIYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
10/21/2011

14:11:47

Standard Lot#: 01011100A1
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	14:12:18
0.100 Cal Check	0.099	14:12:39
Air Blank	0.000	14:13:08

Preventative Maintenance
Performed -- Timer Reset



A handwritten signature in black ink, followed by the date "10/21/2011" written in a similar cursive style.