

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. A. [Signature] # 252 AGENCY Prescott P.D.
DATE 11 15 12 TIME 1018
INTOXILYZER SERIAL # 80-000337 LOCATION P.P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC. into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 252

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
11/15/2012 10:18:29

Standard Lot#: 09312100A1
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	10:19:00
0.100 Cal Check	0.100	10:19:20
Air Blank	0.000	10:19:50

Preventative Maintenance
Performed -- Timer Reset

T. Attenberger #252

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST T. A. ALMAGAL #252 AGENCY PRISON P.D.
DATE 11 15 12 TIME 1020
INTOXILYZER SERIAL # 80-000337 LOCATION P.P.D.

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check,

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

ALL FUNCTIONS WRO CORRECTLY.

SIGNATURE D. H. 252

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
11/15/2012

10:20:46

Standard Lot#: 09312100A1
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP,TEST
DOB: 01/01/1900
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:21:47
Diagnostic Test	Pass	10:22:21
Air Blank	0.000	10:22:48
0.100 Cal Check	0.100	10:23:09
Air Blank	0.000	10:23:38
Subject Test	0.000	10:24:04
Air Blank	0.000	10:24:32
Wait	ABT*	
Air Blank	0.000	10:25:08

*Sequence Aborted

Atty 252

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
11/15/2012

10:25:45

Standard Lot#: 09312100A1
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP,TEST
DOB: 01/01/1900
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	10:26:18
Air Blank	0.000	10:26:46

*Improper Sample



Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
11/15/2012

10:27:11

Standard Lot#: 09312100A1
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1900
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:27:59
Air Blank	0.000	10:28:27

*RFI Detect



A handwritten signature in cursive script, appearing to read 'T. Attenberger', followed by the number '252'.

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
11/15/2012

10:28:48

Standard Lot#: 09312100A1
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1900
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:29:41
Diagnostic Test	Pass	10:30:15
Air Blank	0.000	10:30:42
0.100 Cal Check	0.100	10:31:03
Air Blank	0.000	10:31:33
Subject Test	0.052*	10:34:40
Air Blank	0.000	10:35:12
Air Blank	0.000	10:35:39
Subject Test	XXX**	10:36:07
Air Blank	0.000	10:36:40

*Deficient Sample

**Invalid Sample

Begin new deprivation period



Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. ANTONBERGER #250 AGENCY Prescott P.D.
DATE 11 15 12 TIME 1037
INTOXILYZER SERIAL # 80-000337 LOCATION P.P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution -100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(✓) 4. Air blank completed.
(✓) 5. Calibration check completed. Test results 0. 096 AC.
(✓) 6. Air blank completed.
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE T. Antonberger #250

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
11/15/2012 10:37:19

Standard Lot#: 09312100A1
Last Changed By: T. ATTENBERGER #252

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	10:37:50
0.100 Cal Check	0.096	10:38:11
Air Blank	0.000	10:38:41

Preventative Maintenance
Performed -- Timer Reset

