


EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MEDINA AGENCY PRESOTT
DATE 6/18/13 TIME 1653
INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 099 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

NOTE: DAS NOTIFIED ON 6/14/13 "CAL" CHECK ERROR
DAN SMITH REPORTING OFFICER.

PRESOTT PID CHANGE 5-12-06
SD HAYTEL #127

6/18/13
u

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

06/18/2013

16:53:27

Standard Lot#: 34612100A1

Last Changed By: M. MEDINA #356

QAS: M. MEDINA #356

PRESCOTT PD

Operator: M. MEDINA #356

PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	16:54:12
0.100 Cal Check	0.099	16:54:33
Air Blank	0.000	16:55:03

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K. MEDINA AGENCY PRESCOTT
DATE 6/18/13 TIME 1059
INTOXILYZER SERIAL # 80-000337 LOCATION PRESCOTT P.D.

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

DPS Form Exh G-4 (Rev 05-01)

CONTACTED JOE S. @ DPS CRIME LAB. REQUESTED
I DO A G4 ON INSTRUMENT AFTER "CAC" CHECK ERROR
OBSERVED ON 6/14/13

PRESCOTT PD CHANGE 5 18:06
SD SYSTEM #127

6/18/13
H

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
06/18/2013

16:59:16

Standard Lot#: 34612100A1
Last Changed By: M. MEDINA #356

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP,TEST
DOB: 02/02/2000
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:00:48
Diagnostic Test	Pass	17:01:22
Air Blank	0.000	17:01:49
0.100 Cal Check	0.100	17:02:10
Air Blank	0.000	17:02:40
Subject Test	0.000	17:03:11
Air Blank	0.000	17:03:38
Wait	ABT*	
Air Blank	RFI**	17:04:08

*Sequence Aborted
**RFI Detect

Not a Successfully Completed Test Sequence

6/18/13

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
06/18/2013

17:05:16

Standard Lot#: 34612100A1
Last Changed By: M. MEDINA #356

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP, TEST
DOB: 02/02/2000
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	17:06:06
Air Blank	0.000	17:06:34

*Improper Sample

Not a Successfully Completed Test Sequence

6/18/13
P

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
06/18/2013 17:06:56

Standard Lot#: 34612100A1
Last Changed By: M. MEDINA #356

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP,TEST
DOB: 02/02/2000
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:07:51
Diagnostic Test	Pass	17:08:25
Air Blank	0.000	17:08:52
0.100 Cal Check	0.099	17:09:13
Air Blank	0.000	17:09:43
Subject Test	0.037*	17:12:50
Air Blank	0.000	17:13:22
Air Blank	0.000	17:13:49
Subject Test	XXX**	17:14:14
Air Blank	0.000	17:14:47

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MEDINA AGENCY PRESLOTT P.D.
DATE 6/18/13 TIME _____
INTOXILYZER SERIAL # 80-000337 LOCATION PRESLOTT P.D.

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

PRESLOTT PD CHANGE 5-12-06
SD NAYTEL #127

