

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. Amador #252 AGENCY Prescott P.D.
DATE 10/18/13 TIME 0958
INTOXILYZER SERIAL # 80-000337 LOCATION Prescott P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select (D) for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  #252

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

10/18/2013

09:58:09

Standard Lot#: 34612100A1

Last Changed By: M. MEDINA #356

QAS: T. ATTENBERGER #252

PRESCOTT PD

Operator: T. ATTENBERGER #252

PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	09:58:39
0.100 Cal Check	0.098	09:59:00
Air Blank	0.000	09:59:30

Preventative Maintenance
Performed -- Timer Reset

A handwritten signature in black ink, consisting of a large, stylized initial 'A' followed by a series of loops and a final flourish.

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST T. Anderson #352 AGENCY Prescott PIA
DATE 10/18/13 TIME 1004
INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT PIA

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

ALL FUNCTIONING CORRECTLY.

SIGNATURE

T. Anderson #352

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
10/18/2013

10:04:31

Standard Lot#: 34612100A1
Last Changed By: M. MEDINA #356

QAS: T. ATTENBERGER #252
PRESCOTT PD


Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1900
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:05:39
Diagnostic Test	Pass	10:06:13
Air Blank	0.000	10:06:40
0.100 Cal Check	0.097	10:07:01
Air Blank	0.000	10:07:31
Subject Test	0.000	10:07:53
Air Blank	0.000	10:08:21
Wait	ABT*	
Air Blank	0.000	10:08:58

*Sequence Aborted

A handwritten signature in black ink, appearing to be 'T. Attenberger', followed by a horizontal line.

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
10/18/2013

10:09:40

Standard Lot#: 34612100A1
Last Changed By: M. MEDINA #356

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1900
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	10:10:19
Air Blank	0.000	10:10:47

*Improper Sample



Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

10/18/2013

10:11:15

Standard Lot#: 34612100A1

Last Changed By: M. MEDINA

#356

QAS: T. ATTENBERGER

#252

PRESCOTT PD

Operator: T. ATTENBERGER

#252

PRESCOTT PD

Subject: SQAP, TEST

DOB: 01/01/1900

Sex: M

Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:12:05
Air Blank	0.000	10:12:34

*RFI Detect



Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
10/18/2013

10:13:03

Standard Lot#: 34612100A1
Last Changed By: M. MEDINA #356

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Subject: SQAP, TEST
DOB: 01/01/1900
Sex: M Weight: 300

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:13:59
Diagnostic Test	Pass	10:14:33
Air Blank	0.000	10:15:00
0.100 Cal Check	0.097	10:15:21
Air Blank	0.000	10:15:51
Subject Test	0.047*	10:18:58
Air Blank	0.000	10:19:30
Air Blank	0.000	10:19:58
Subject Test	XXX**	10:20:34
Air Blank	0.000	10:21:07

*Deficient Sample

**Invalid Sample

Begin new deprivation period



Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST T. Amberg #252 AGENCY Prison PD
DATE 10/18/13 TIME 1021
INTOXILYZER SERIAL # 80-000337 LOCATION Prison PD

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select (D) for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 096 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 252

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
10/18/2013 10:21:43

Standard Lot#: 34612100A1
Last Changed By: M. MEDINA #356

QAS: T. ATTENBERGER #252
PRESCOTT PD

Operator: T. ATTENBERGER #252
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	10:22:13
0.100 Cal Check	0.096	10:22:34
Air Blank	0.000	10:23:03

Preventative Maintenance
Performed -- Timer Reset

