

County Attorney



ARIZONA DEPARTMENT OF PUBLIC SAFETY

INTOXILYZER MAINTENANCE REPORT

AGENCY Prescott Police Department
Prescott, AZ 86303

Page 1 of 1

OFFICER ATTENBERGER, #252

DATE January 17, 2014

AF 1/21/14

Dissemination is restricted to:
Criminal Justice Agencies ONLY

PRESCOTT POLICE DEPARTMENT
REL TO: COUNTY ATTORNEY

INSTRUMENT:

80000337

MAINTENANCE REQUESTED:

Repair and Preventative Maintenance

MAINTENANCE PERFORMED:

Parts Replaced:

Printer

Preventative Maintenance Performed:

Inspection, Calibration, Verification, Adjustment

A handwritten signature in cursive script, appearing to read "Joseph M Slowinski".

JOSEPH M SLOWINSKI, #6101
Electronics Technician
Central Regional Crime Laboratory
P. O. Box 6638
Phoenix, AZ 85005
(602) 223-2394



EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MEDINA AGENCY PRESOTT P.D.
DATE 1/20/14 TIME 1817
INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 097 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

RETURNED TO SERVICE

M. MEDINA
1/20/14

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
01/20/2014 18:17:04

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	18:17:35
0.100 Cal Check	0.097	18:17:56
Air Blank	0.000	18:18:26

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MEDINA AGENCY PRESOTT P.D.

DATE 1/20/14 TIME 1819

INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

MISSED AIR BLANK PUSHED/KEYED UP RADIO
SECOND TEST INITIATED FOR RET

SIGNATURE [Signature]

M. MEDINA
1/20/14

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
01/20/2014 18:19:15

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP, TEST
DOB: 02/02/2000
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:20:39
Diagnostic Test	Pass	18:21:13
Air Blank	0.000	18:21:41
0.100 Cal Check	0.098	18:22:02
Air Blank	0.000	18:22:32
Subject Test	0.000	18:22:52
Air Blank	0.000	18:23:20
Wait	ABT*	
Air Blank	IPS**	18:23:41

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

M. MEDINA 356
1/20/14

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

01/20/2014

18:24:41

Standard Lot#: 25913100A2

Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356

PRESCOTT PD

Operator: M. MEDINA #356

PRESCOTT PD

Subject: SQAP, TEST

DOB: 02/02/2000

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:25:40
Diagnostic Test	Fail*	18:25:45
Air Blank	RFI**	18:25:52

* Voltage/Current Test

**RFI Detect

Not a Successfully Completed Test Sequence

M. MEDINA 3IL
4/20/14

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
01/20/2014 18:26:37

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP,TEST
DOB: 02/02/2000
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	18:27:19
Air Blank	0.000	18:27:48

*RFI Detect

Not a Successfully Completed Test Sequence

M. MEDINA
1/20/14

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
01/20/2014 18:30:46

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP, TEST
DOB: 02/02/2000
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:31:44
Diagnostic Test	Pass	18:32:18
Air Blank	0.000	18:32:46
0.100 Cal Check	0.097	18:33:06
Air Blank	0.000	18:33:36
Subject Test	0.023*	18:36:44
Air Blank	0.000	18:37:14
Air Blank	0.000	18:37:41
Subject Test	XXX**	18:38:02
Air Blank	0.000	18:38:35

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MEDINA AGENCY PRESOTT P.D.
DATE 1/20/14 TIME _____
INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 096 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

M. MEDINA
1/20/14

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
01/20/2014 18:39:13

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	18:39:44
0.100 Cal Check	0.096	18:40:05
Air Blank	0.000	18:40:35

Preventative Maintenance
Performed -- Timer Reset