

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST MATT MEDINA AGENCY PRESLOTT P.D.
DATE 5/1/14 TIME 1314
INTOXILYZER SERIAL # 80 000337 LOCATION PRESLOTT P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.097 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

Prescott PD CHANGE 5-12-06
ID NO. 052 8127

mu
5/1/14

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

05/01/2014

13:14:37

Standard Lot#: 25913100A2

Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356

PRESCOTT PD

Operator: M. MEDINA #356

PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	13:15:41
0.100 Cal Check	0.097	13:16:01
Air Blank	0.000	13:16:31

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST MATT MEDINA AGENCY PRESOTT P.D.
DATE 5/1/14 TIME 1317
INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT P.D.

(1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (1. Clock time check.
(2. Date check.

OPERATIONAL TESTS

- (1. Alcohol-free subject test result 0. 000 AC.
(2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

m
5/1/14

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
05/01/2014

13:17:34

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP, TEST
DOB: 02/02/2000
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:18:33
Diagnostic Test	Pass	13:19:07
Air Blank	0.000	13:19:34
0.100 Cal Check	0.097	13:19:55
Air Blank	0.000	13:20:25
Subject Test	0.000	13:20:54
Air Blank	0.000	13:21:22
Wait	ABT*	
Air Blank	RFI**	13:21:51

*Sequence Aborted
**RFI Detect

Not a Successfully Completed Test Sequence

M
5/1/14

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
05/01/2014 13:24:46

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP, TEST
DOB: 02/02/2000
Sex: M . Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	13:25:26
Air Blank	0.000	13:25:55

*Improper Sample

Not a Successfully Completed Test Sequence

5/1/14

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
05/01/2014 13:26:30

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP, TEST
DOB: 02/02/2000
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:27:24
Diagnostic Test	Pass	13:27:58
Air Blank	0.000	13:28:26
0.100 Cal Check	0.096	13:28:47
Air Blank	0.000	13:29:16
Subject Test	0.033*	13:32:24
Air Blank	0.000	13:32:55
Air Blank	0.000	13:33:23
Subject Test	XXX**	13:33:54
Air Blank	0.000	13:34:27

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST MATT MEDINA AGENCY PRESOTT P.D.
DATE 5/1/14 TIME _____
INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 096 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

PRESOTT P.D. CHANGE 5-18-06

INSTRUMENT # 127

5/1/14

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
05/01/2014 13:35:42

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	13:36:13
0.100 Cal Check	0.096	13:36:34
Air Blank	0.000	13:37:04

Preventative Maintenance
Performed -- Timer Reset