

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MEDINA AGENCY PRESOTT
DATE 6/26/14 TIME 1520
INTOXILYZER SERIAL # 80-000 337 LOCATION PRESOTT P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 096 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

6/20/17

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
06/26/2014 15:20:56

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	15:21:26
0.100 Cal Check	0.096	15:21:47
Air Blank	0.000	15:22:17

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MEDINA AGENCY PRESOTT
DATE 6/20/14 TIME 1524
INTOXILYZER SERIAL # 80-000337 LOCATION PRESOTT P.D.

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 094 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

TAKEN OUT OF SERVICE / NEW INSTRUMENT SQAP COMPLETED
BY T. ATTENDEGER 252

SIGNATURE 

PRESOTT PD CHANGE 5 18:06
TO NUMBER 252

6/26/14
✓

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
06/26/2014

15:24:25

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP, TEST
DOB: 02/02/2000
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:25:19
Diagnostic Test	Pass	15:25:53
Air Blank	0.000	15:26:21
0.100 Cal Check	0.096	15:26:42
Air Blank	0.000	15:27:12
Subject Test	0.000	15:27:35
Air Blank	0.000	15:28:02
Wait	ABT*	
Air Blank	RFI**	15:28:25

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

6/26/14
m

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

06/26/2014

15:28:37

Standard Lot#: 25913100A2

Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356

PRESCOTT PD

Operator: M. MEDINA #356

PRESCOTT PD

Subject: SQAP,TEST

DOB: 02/02/2000

Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	15:29:26
Air Blank	0.000	15:29:54

*Improper Sample

Not a Successfully Completed Test Sequence

6/26/14
✓

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
06/26/2014 15:30:12

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Subject: SQAP, TEST
DOB: 02/02/2000
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:31:25
Diagnostic Test	Pass	15:31:58
Air Blank	0.000	15:32:26
0.100 Cal Check	0.094	15:32:47
Air Blank	0.000	15:33:17
Subject Test	0.033*	15:36:24
Air Blank	0.000	15:36:56
Air Blank	0.000	15:37:23
Subject Test	XXX**	15:37:50
Air Blank	0.000	15:38:24

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MEDINA AGENCY PRESLWIT
DATE 2/26/14 TIME 1539
INTOXILYZER SERIAL # 80-000337 LOCATION PRESLWIT P.D.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 094 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

Prescott P.D. CHANGE 5-18-06

80-000337

6/26/14

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
06/26/2014 15:39:13

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: M. MEDINA #356
PRESCOTT PD

Operator: M. MEDINA #356
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	15:39:44
0.100 Cal Check	0.094	15:40:05
Air Blank	0.000	15:40:35

Preventative Maintenance
Performed -- Timer Reset