

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST STEAN HILTON AGENCY PRESCOTT P.D.
DATE 3/11/2015 TIME 1509
INTOXILYZER SERIAL # 80-000337 LOCATION 222 S. MARINA ST.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE S. Hilton

DPS Form Exh G-2 (Rev 05-01)

SMA 372

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
03/11/2015 15:13:28

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	15:14:00
0.100 Cal Check	0.098	15:14:21
Air Blank	0.000	15:14:51

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST STEVE HESTON AGENCY PRESOTT P.D.
DATE 3/11/2015 TIME 1515
INTOXILYZER SERIAL # 80-000337 LOCATION 222 S. MARINA ST.

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 100 AC
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

operating properly

SIGNATURE S. HESTON

SMA
SR

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

03/11/2015

15:17:17

Standard Lot#: 25913100A2

Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372

PRESCOTT PD

Operator: S. HILTON #372

PRESCOTT PD

Subject: SQAP,TEST

DOB: 12/19/1973

Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:18:40
Diagnostic Test	Pass	15:19:14
Air Blank	0.000	15:19:42
0.100 Cal Check	0.097	15:20:03
Air Blank	0.000	15:20:33
Subject Test	0.000	15:20:51
Air Blank	ABT*	15:21:18
Air Blank	0.000	15:21:46

*Sequence Aborted

Not a Successfully Completed Test Sequence

SMA 372

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
03/11/2015

15:22:12

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Subject: SQAP, TEST
DOB: 12/19/1973
Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	15:23:02
Air Blank	0.000	15:23:30

*Improper Sample

Not a Successfully Completed Test Sequence

SMA 372

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
03/11/2015

15:24:04

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Subject: SQAP,TEST
DOB: 12/19/1973
Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	15:24:48
Air Blank	0.000	15:25:16

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
03/11/2015

SMA 372

15:25:44

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Subject: SQAP, TEST
DOB: 12/10/1973
Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:26:59
Diagnostic Test	Pass	15:27:33
Air Blank	0.000	15:28:01
0.100 Cal Check	0.097	15:28:22
Air Blank	0.000	15:28:52
Subject Test	0.148*	15:31:59
Air Blank	0.000	15:32:43
Air Blank	0.000	15:33:10
Subject Test	XXX**	15:33:30
Air Blank	0.000	15:34:06

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Stacy Horton AGENCY Prescott P.D.
DATE 3/11/2015 TIME 1535
INTOXILYZER SERIAL # 80-000337 LOCATION 222 S. MARINA ST.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution _____ AC.
OR
Pour a standard alcohol concentration solution _____ AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select D for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 097 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

DPS Form Exh G-2 (Rev 05-01)

SMA 3h

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
03/11/2015

15:35:15

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	15:35:46
0.100 Cal Check	0.097	15:36:07
Air Blank	0.000	15:36:37

Preventative Maintenance
Performed -- Timer Reset