

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST S. HILTON AGENCY Prescott P.D.
DATE 4/16/15 TIME 1911
INTOXILYZER SERIAL # 80-000337 LOCATION 222 S. MARINA ST

- (S)* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (S)* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(S) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(S) 4. Air blank completed.
(S) 5. Calibration check completed. Test results 0. 098 AC.
(S) 6. Air blank completed.
(S) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE S. Hilton

INTOXILYZER 8000

Location: PRESCOTT P.D

Serial Number: 80-000337

Core Version: 8105.48

04/10/2015

19:15:09

Standard Lot#: 25913100A2

Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372

PRESCOTT PD

Operator: S. HILTON #372

PRESCOTT PD

Test g/210L Time

Air Blank 0.000 19:15:40
0.100 Cal Check 0.098 19:16:01
Air Blank 0.000 19:16:31

Preventative Maintenance
Performed -- Timer Reset

SMA 372

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST S. HEITON AGENCY PRESCOTT P.D.
DATE 04/20/15 TIME 1918
INTOXILYZER SERIAL # 80-000337 LOCATION 222 S. MARION ST.

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 038 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

Functioning Properly

SIGNATURE S. HEITON

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
04/10/2015

19:18:50

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Subject: SQAP, TEST
DOB: 12/19/1973
Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:20:13
Diagnostic Test	Pass	19:20:47
Air Blank	0.000	19:21:15
0.100 Cal Check	0.098	19:21:36
Air Blank	0.000	19:22:06
Subject Test	0.000	19:22:22
Air Blank	ABT*	19:22:50
Air Blank	0.000	19:23:18

*Sequence Aborted



Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
04/10/2015

19:24:00

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Subject: SQAP, TEST
DOB: 12/19/1973
Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	19:25:10
Air Blank	0.000	19:25:38

*Improper Sample

SMA 772

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
04/10/2015

19:27:13

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Subject: SQAP, TEST
DOB: 12/19/1973
Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	19:27:55
Air Blank	0.000	19:28:23

*RFI Detect

SMA 372

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
04/10/2015

19:29:06

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Subject: SQAP, TEST
DOB: 12/19/1973
Sex: M Weight: 175

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	19:30:07
Diagnostic Test	Pass	19:30:41
Air Blank	0.000	19:31:08
0.100 Cal Check	0.098	19:31:29
Air Blank	0.000	19:31:59
Subject Test	0.138*	19:35:06
Air Blank	0.000	19:35:43
Air Blank	0.000	19:36:10
Subject Test	XXX**	19:36:36
Air Blank	0.000	19:37:13

*Deficient Sample

**Invalid Sample

Begin new deprivation period



Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST S. HILTON AGENCY Prescott P.D.

DATE 4/10/15 TIME 1938

INTOXILYZER SERIAL # 80-000337 LOCATION 222 S. MARINA ST.

(S) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR -

Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(S) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(S) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(S) 4. Air blank completed.

(S) 5. Calibration check completed. Test results 0. 098 AC.

(S) 6. Air blank completed.

(S) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE S. HILTON

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
04/10/2015 19:38:10

Standard Lot#: 25913100A2
Last Changed By: T. ATTENBERGER #252

QAS: S. HILTON #372
PRESCOTT PD

Operator: S. HILTON #372
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	19:38:44
0.100 Cal Check	0.098	19:39:05
Air Blank	0.000	19:39:35

Preventative Maintenance
Performed -- Timer Reset

SMA 372