

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. BARDON AGENCY PRESCOTT P.D.

DATE AUG 27, 2018 TIME 1535

INTOXILYZER SERIAL # 80-000337 LOCATION PRESCOTT P.D.

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record: Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
08/27/2018

15:34:31

Standard Lot#: 22817100A3
Last Changed By: D. BARDON #351

QAS: D. BARDON #351
PRESCOTT PD

Operator: D. BARDON #351
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	15:35:16
0.100 Cal Check	0.098	15:35:36
Air Blank	0.000	15:36:07

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST D. BARDON AGENCY Prescott P.D.
DATE Aug 27, 2018 TIME 1538
INTOXILYZER SERIAL # 90-000337 LOCATION Prescott P.D.

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE [Signature]

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
08/27/2018

15:37:26

Standard Lot#: 22817100A3
Last Changed By: D. BARDON #351

QAS: D. BARDON #351
PRESCOTT PD

Operator: D. BARDON #351
PRESCOTT PD

Subject: SQAP, TEST
DOB: 09/09/1999
Sex: M Weight: 210

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:38:43
Diagnostic Test	Pass	15:39:17
Air Blank	0.000	15:39:45
0.100 Cal Check	0.098	15:40:06
Air Blank	0.000	15:40:36
Subject Test	0.000	15:40:54
Air Blank	ABT*	15:41:00
Air Blank	0.000	15:41:28

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
08/27/2018

15:42:25

Standard Lot#: 22817100A3
Last Changed By: D. BARDON #351

QAS: D. BARDON #351
PRESCOTT PD

Operator: D. BARDON #351
PRESCOTT PD

Subject: SQAP,TEST
DOB: 09/09/1999
Sex: M Weight: 210

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	15:43:06
Air Blank	RFI**	15:43:20

*Improper Sample
**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
08/27/2018

15:44:06

Standard Lot#: 22817100A3
Last Changed By: D. BARDON #351

QAS: D. BARDON #351
PRESCOTT PD

Operator: D. BARDON #351
PRESCOTT PD

Subject: SQAP,TEST
DOB: 09/09/1999
Sex: M Weight: 210

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:44:58
Diagnostic Test	Pass	15:45:32
Air Blank	0.000	15:46:00
0.100 Cal Check	0.097	15:46:20
Air Blank	0.000	15:46:51
Subject Test	0.029*	15:49:58
Air Blank	0.000	15:50:30
Air Blank	0.000	15:50:57
Subject Test	XXX**	15:51:18
Air Blank	0.000	15:51:52

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. BARDON AGENCY Prescott P.D.
DATE AUG, 27, 2018 TIME 1555
INTOXILYZER SERIAL # 80-000337 LOCATION Prescott P.D.

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record; Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000
Location: PRESCOTT P.D
Serial Number: 80-000337
Core Version: 8105.48
08/27/2018

15:54:49

Standard Lot#: 22817100A3
Last Changed By: D. BARDON #351

QAS: D. BARDON #351
PRESCOTT PD

Operator: D. BARDON #351
PRESCOTT PD

Test	g/210L	Time
Air Blank	0.000	15:55:20
0.100 Cal Check	0.097	15:55:40
Air Blank	0.000	15:56:11

Preventative Maintenance
Performed -- Timer Reset