

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST S.WILLADSEN #886 AGENCY SEDONA P.D.  
DATE 01/03/2026 TIME 1027  
INTOXILYZER SERIAL # 90-003521

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample - Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0.099 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO


I hereby certify that the above and foregoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 9000 reference therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety

Date this 03 day of JANUARY 2026

  
S. WILLADSEN  
Quality Assurance Specialist

COMMENTS:

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE  S.WILLADSEN #886

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

SEDONA PD

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-003521  
Software Version: 9439.01.00  
Analytical Report Number: 35210103260141

QAS: WILLADSEN, STEVEN  
QAS Permit #: 30140  
Agency: SEDONA PD  
Last 31-Day Check: 12/06/2025  
Last Annual Maintenance: 04/18/2025

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 01/03/2026  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 10:00  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: WILLADSEN, STEVEN  
Agency: SEDONA PD

Permit #: 30140

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 10124100A1  
Expiration Date: 06/05/2026  
Bottle #: 046  
Last Changed By: WILLADSEN, STEVEN

Test	g/210L	Time
Air Blank	0.000	10:27:11
Diagnostic Test	PASS	10:27:47
Air Blank	0.000	10:28:24
Calibration Chk	0.099	10:28:45
Air Blank	0.000	10:29:25
Subject Test 1	DEF*	10:32:54
Air Blank	0.000	10:33:39
Air Blank	0.000	10:34:19
Subject Test 2	0.000	10:35:02
Air Blank	0.000	10:35:43
Wait		10:39:47
Air Blank	0.000	10:40:24
Subject Test 3	INV**	10:41:01
Air Blank	RFI***	10:41:14

Permit #: 30140

## RESULTS

## EXCEPTION MESSAGES

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

I hereby certify that the above and foregoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 9000 reference therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety.

## OPERATOR COMMENTS

Date this 03 day of JANUARY 2026

  
S. WILLADSEN  
Quality Assurance Specialist

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.