

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J Young #1110 AGENCY Sedona PD
DATE 2/1/26 TIME 0750
INTOXILYZER SERIAL # 90-003521

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0.00 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0.100 AC
6. Air Blanks Completed
7. Timer Reset

I hereby certify that the above and foregoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 9000 reference therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety

Date this 1st day of February 2026

J Young #1110
Quality Assurance Specialist

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS: Annual Maint Expires 04/18/26

SIGNATURE J Young #1110

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

SEDONA PD

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003521
Software Version: 9439.01.00
Analytical Report Number: 35210201260146

QAS: WILLADSEN, STEVEN
QAS Permit #: 30140
Agency: SEDONA PD
Last 31-Day Check: 01/03/2026
Last Annual Maintenance: 04/18/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 02/01/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 07:35
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: YOUNG, JASON
Agency: SEDONA PD

Permit #: 30377

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 10124100A1
Expiration Date: 06/05/2026
Bottle #: 046
Last Changed By: WILLADSEN, STEVEN

Permit #: 30140

Test	g/210L	Time
Air Blank	0.000	07:50:25
Diagnostic Test	PASS	07:51:00
Air Blank	0.000	07:51:37
Calibration Chk	0.100	07:51:59
Air Blank	0.000	07:52:39
Subject Test 1	DEF*	07:55:54
Air Blank	0.000	07:56:37
Air Blank	0.000	07:57:17
Subject Test 2	0.000	07:58:06
Air Blank	0.000	07:58:56
Wait		08:02:59
Air Blank	0.000	08:03:36
Subject Test 3	INV**	08:04:07
Air Blank	RFI***	08:04:20

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

I hereby certify that the above and foregoing is a true and correct copy of the record of periodic maintenance for the intoxilyzer 9000 reference therein maintained by the Sedona Police Department pursuant to the requirements of the Arizona Department of Public Safety

Date this 1st day of February 2026

J Young #1110
Quality Assurance Specialist

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.