

25-05880

COPY

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST E. McLAEN AGENCY YCSO
DATE 10/22/25 TIME 23:15
INTOXILYZER SERIAL # 90-003588

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0.101 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS: ADJUSTED TIME 11 MINUTES,
OPERATOR 30404
25-050880

SIGNATURE 

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003588
Software Version: 9439.01.00
Analytical Report Number: 35881022250031

QAS: BLAKLEY, SCOTT
QAS Permit #: 30134
Agency: YAVAPAI CSO
Last 31-Day Check: 09/01/2025
Last Annual Maintenance: 12/09/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 10/22/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 22:45
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MCLAIN, EVAN
Agency: YAVAPAI CSO

Permit #: 30404

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-40318349
Expiration Date: 01/24/2027
Bottle #: 16
Last Changed By: MCLAIN, EVAN

Permit #: 30404

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

30404 OPERATOR ID. ADJUSTED TIME DIFF OF 11MIN

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.*

Test	g/210L	Time
Air Blank	0.000	23:03:04
Diagnostic Test	PASS	23:03:39
Air Blank	0.000	23:04:16
Calibration Chk	0.101	23:04:38
Air Blank	0.000	23:05:18
Subject Test 1	DEF*	23:08:35
Air Blank	0.000	23:09:20
Air Blank	0.000	23:10:00
Subject Test 2	0.000	23:10:29
Air Blank	0.000	23:11:10
Wait		23:15:14
Air Blank	0.000	23:15:51
Subject Test 3	INV**	23:16:33
Air Blank	RFI***	23:16:46

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence