

25-050875

COPY

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST E MCLAEN AGENCY YCSO
DATE 10/22/25 TIME 2130
INTOXILYZER SERIAL # 90-003589

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0. 098 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS: MACHINE DUE FOR ~~REPAIR~~ ANNUAL
ON 10-30-25 DR 25-050875

SIGNATURE E MCLAEN

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402972207
Expiration Date: 02/20/2026
Bottle #: 21
Last Changed By:
BLAKLEY, SCOTT

TEST	RESULT	TIME
Calibration Chk	0.098	21:16:11
Air Blank	0.000	21:16:50
Subject Test 1	DEF*	21:20:19
Air Blank	0.000	21:21:07
Air Blank	0.000	21:21:47
Subject Test 2	0.000	21:24:00
Air Blank	0.000	21:24:40
Wait		21:28:44
Air Blank	0.000	21:29:21
Subject Test 3	INV**	21:30:24
Air Blank	RFI***	21:30:35

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-003589
Software Version: 9439.01.00
Analytical Report Number: 35891022250034
QAS: BLAKLEY, SCOTT
QAS Permit #: 30134
Agency: YAVAPAI CSO
Last 31-Day Check: 09/01/2025
Last Annual Maintenance: 10/30/2024

SUBJECT INFORMATION

Test Date: 10/22/2025
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 20:30
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MCLAIN, EVAN
Agency: YAVAPAI CSO
Permit #: 30404