

26-014086

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST R. CEDRO AGENCY YAVAPAI COUNTY S.O.
DATE 03/20/2026 TIME 1734
INTOXILYZER SERIAL # 90-004105

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration _____ AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 100 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0. _____ AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO _____

COMMENTS:

RETURNED TO SERVICE.

SIGNATURE

R. CEDRO #13294

26-014086

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004105
Software Version: 9439.01.00
Analytical Report Number: 41050320260080

QAS: BOLING, LUKE
QAS Permit #: 30212
Agency: YAVAPAI CSO
Last 31-Day Check: 01/22/2026
Last Annual Maintenance: 10/03/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 03/20/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 17:15
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: CEDRO, RYAN
Agency: YAVAPAI CSO

Permit #: 30397

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-403183349
Expiration Date: 01/24/2027
Bottle #: 30
Last Changed By: KAUFMAN, JASON

Permit #: 30402

Test	g/210L	Time
Air Blank	0.000	17:34:01
Diagnostic Test	PASS	17:34:36
Air Blank	0.000	17:35:13
Calibration Chk	0.099	17:35:35
Air Blank	0.000	17:36:15
Subject Test 1	DEF*	17:39:29
Air Blank	0.000	17:40:10
Air Blank	0.000	17:40:52
Subject Test 2	0.000	17:41:27
Air Blank	0.000	17:42:14
Wait		17:46:17
Air Blank	0.000	17:46:54
Subject Test 3	INV**	17:47:50
Air Blank	RFI***	17:48:03

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTSEXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.