

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENTANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST R. Rivera AGENCY Yuma CO  
DATE 11-11-25 TIME 2:07  
INTOXILYZER SERIAL # 90-004106

1. Ensure that gas tank is attached and contains a standard alcohol concentration .160 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample - Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- 5. Standard Calibration Check:  
0.097 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE  6354

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-004106  
Software Version: 9439.01.00  
Analytical Report Number: 41061111250056

QAS: RIVERA, REYMUNDO  
QAS Permit #: 30268  
Agency: YAVAPAI CSO  
Last 31-Day Check: 10/08/2025  
Last Annual Maintenance: 07/31/2025

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 11/11/2025  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 20:30  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: RIVERA, REYMUNDO  
Agency: YAVAPAI CSO

Permit #: 30268  
STANDARD INFORMATION  
Standard Value: 0.100  
Standard Lot #: 302-40318334932  
Expiration Date: 1/24/2027  
Bottle #: 31  
Last Changed By: KAUFMAN, JASON

Test	g/210L	Time
Air Blank	0.000	21:07:53
Diagnostic Test	PASS	21:08:29
Air Blank	0.000	21:09:05
Calibration Chk	0.100	21:09:27
Air Blank	0.000	21:10:06
Subject Test 1	DEF*	21:13:20
Air Blank	0.000	21:16:17
Air Blank	0.000	21:16:58
Subject Test 2	0.000	21:17:27
Air Blank	0.000	21:18:33
Wait		21:22:36
Air Blank	0.000	21:23:13
Subject Test 3	INV**	21:23:47
Air Blank	RFI***	21:24:04

Permit #: 30402

RESULTS

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

## EXCEPTION MESSAGES

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

YARNELL

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*

25-054221

Intoxilyzer - Alcohol Analyzer  
Model 9000

SN: 90-004106

Test Date: 11/11/2025

Test Time: 21:25

Location:

Software Version: 9439.01.00

Standard Lot #: 302-40318334932

Standard Value: 0.100

Expiration Date: 01/24/2027

Bottle #: 31

Last Changed By: KAUFMAN, JASON

Permit #: 30402

Operator: RIVERA, REYMUNDO

Permit #: 30268

Agency: YAVAPAI CSO

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	21:25:28
Calibration Chk	0.097	21:25:50
Air Blank	0.000	21:26:30

No RFI Detected