

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENTANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST BOLING, L #11323 AGENCY YAVAPAI COUNTY SHERIFF'S OFFICE
DATE 11/06/2025 TIME 2315
INTOXILYZER SERIAL # 90004107

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0.095 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE Luke Boling #11323

25-053336

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004107
Software Version: 9439.01.00
Analytical Report Number: 41071106250066

QAS: BOLING, LUKE
QAS Permit #: 30212
Agency: YAVAPAI CSO
Last 31-Day Check: 09/22/2025
Last Annual Maintenance: 10/03/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 11/06/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 22:55
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: BOLING, LUKE
Agency: YAVAPAI CSO

Permit #: 30212

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402972207
Expiration Date: 05/02/2026
Bottle #: 17
Last Changed By: BOLING, LUKE

Permit #: 30212

Test	g/210L	Time
Air Blank	0.000	23:15:25
Diagnostic Test	PASS	23:16:01
Air Blank	0.000	23:16:38
Calibration Chk	0.095	23:16:59
Air Blank	0.000	23:17:39
Subject Test 1	DEF*	23:21:07
Air Blank	0.000	23:21:51
Air Blank	0.000	23:22:31
Subject Test 2	0.000	23:22:56
Air Blank	0.000	23:23:39
Wait		23:27:42
Air Blank	0.000	23:28:19
Subject Test 3	INV**	23:29:29
Air Blank	RFI***	23:29:40

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.*