

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENTANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST BOLING, L #11323 AGENCY YAVAPAI COUNTY SHERIFF'S OFFICE
DATE 12/23/2025 TIME 2151
INTOXILYZER SERIAL # 90004107

- 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0.096 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES X NO

COMMENTS:

SIGNATURE *Luke Boling* #11323

25-061629

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004107
Software Version: 9439.01.00
Analytical Report Number: 41071223250068

QAS: BOLING, LUKE
QAS Permit #: 30212
Agency: YAVAPAI CSO
Last 31-Day Check: 11/06/2025
Last Annual Maintenance: 10/03/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 12/23/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 21:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: BOLING, LUKE
Agency: YAVAPAI CSO

Permit #: 30212

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402972207
Expiration Date: 05/02/2026
Bottle #: 17
Last Changed By: BOLING, LUKE

Permit #: 30212

Test	g/210L	Time
Air Blank	0.000	21:51:39
Diagnostic Test	PASS	21:52:14
Air Blank	0.000	21:52:51
Calibration Chk	0.096	21:53:13
Air Blank	0.000	21:53:52
Subject Test 1	DEF*	21:57:18
Air Blank	0.000	21:58:01
Air Blank	0.000	21:58:43
Subject Test 2	0.000	21:59:25
Air Blank	0.000	22:00:08
Wait		22:04:12
Air Blank	0.000	22:04:49
Subject Test 3	INV**	22:05:38
Air Blank	RFI***	22:05:51

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.*