

## EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

## PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST BOLING, L #11323 AGENCY YAVAPAI COUNTY SHERIFF'S OFFICE  
 DATE 02/16/2026 TIME 1904  
 INTOXILYZER SERIAL # 90004107

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

## DIAGNOSTIC TESTS

1. Clock time check  
 2. Date check

## OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)  
 Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
 Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):  
 RFI Detect printed
5. Standard Calibration Check:  
0.094 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

## COMMENTS:

---



---



---

SIGNATURE Luke Boling #11323

## ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
 Serial Number: 90-004107  
 Software Version: 9439.01.00  
 Analytical Report Number: 41070216260069

QAS: BOLING, LUKE  
 QAS Permit #: 30212  
 Agency: YAVAPAI CSO  
 Last 31-Day Check: 12/23/2025  
 Last Annual Maintenance: 10/03/2025

SUBJECT INFORMATION

Name: 31-Day Check  
 Test Date: 02/16/2026  
 Driver's License #:  
 Gender:  
 Date of Birth:  
 Age:

Weight:  
 State of Issue:  
 Driver's License Expiration:  
 Deprivation Start Time: 18:45  
 15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: BOLING, LUKE  
 Agency: YAVAPAI CSO

Permit #: 30212

STANDARD INFORMATION

Standard Value: 0.100  
 Standard Lot #: 302-402972207  
 Expiration Date: 05/02/2026  
 Bottle #: 17  
 Last Changed By: BOLING, LUKE

Permit #: 30212

Test	g/210L	Time
Air Blank	0.000	19:04:12
Diagnostic Test	PASS	19:04:47
Air Blank	0.000	19:05:24
Calibration Chk	0.094	19:05:46
Air Blank	0.000	19:06:25
Subject Test 1	DEF*	19:09:43
Air Blank	0.000	19:10:26
Air Blank	0.000	19:11:06
Subject Test 2	0.000	19:11:35
Air Blank	0.000	19:12:16
Wait		19:16:19
Air Blank	0.000	19:16:56
Subject Test 3	INV**	19:17:37
Air Blank	RFI***	19:17:50

RESULTSEXCEPTION MESSAGES

\*Deficient Sample  
 \*\*Invalid Sample - Begin new deprivation period  
 \*\*\*RFI Detect

\* Deficient Sample  
 \*\* Invalid Sample - Begin new deprivation period  
 \*\*\* RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.  
 Consecutive subject test results must not differ by more than 0.020 g/210L.  
 Standard check results must be  $\pm 10\%$ .