

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST BOLING, L #11323 AGENCY YAVAPAI COUNTY SHERIFF'S OFFICE
DATE 04/20/2026 TIME 1606
INTOXILYZER SERIAL # 90004107

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0.092 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES X NO

COMMENTS:

SIGNATURE

Luke Boling #11323

26-019914

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004107
Software Version: 9439.01.00
Analytical Report Number: 41070420260070

QAS: BOLING, LUKE
QAS Permit #: 30212
Agency: YAVAPAI CSO
Last 31-Day Check: 02/16/2026
Last Annual Maintenance: 10/03/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 04/20/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 15:45
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: BOLING, LUKE
Agency: YAVAPAI CSO

Permit #: 30212

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302-402972207
Expiration Date: 05/02/2026
Bottle #: 17
Last Changed By: BOLING, LUKE

Permit #: 30212

RESULTS

Test	g/210L	Time
Air Blank	0.000	16:06:08
Diagnostic Test	PASS	16:06:43
Air Blank	0.000	16:07:20
Calibration Chk	0.092	16:07:42
Air Blank	0.000	16:08:21
Subject Test 1	DEF*	16:11:36
Air Blank	0.000	16:12:20
Air Blank	0.000	16:13:01
Subject Test 2	0.000	16:13:30
Air Blank	0.000	16:14:13
Wait		16:18:17
Air Blank	0.000	16:18:54
Subject Test 3	INV**	16:20:37
Air Blank	RFI***	16:20:50

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.