

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST KAUFMAN 14008 AGENCY YCSO
DATE 10/15/25 TIME 2212 HRS
INTOXILYZER SERIAL # 90-004108

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0. 097 AC
6. Air Blanks Completed
7. Timer Reset


Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

ETOH TANK 600 PSI
D/R# 25-049683

SIGNATURE



ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004108
Software Version: 9439.01.00
Analytical Report Number: 41081015250162

QAS: MCLAIN, EVAN
QAS Permit #: 30404
Agency: YAVAPAI CSO
Last 31-Day Check: 09/12/2025
Last Annual Maintenance: 10/30/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 10/15/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 18:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: KAUFMAN, JASON
Agency: YAVAPAI CSO

Permit #: 30402
STANDARD INFORMATION
Standard Value: 0.100
Standard Lot #: 302403183349
Expiration Date: 01/24/2027
Bottle #: 13
Last Changed By: BLAKLEY, SCOTT

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	22:12:36
Diagnostic Test	PASS	22:13:12
Air Blank	0.000	22:13:49
Calibration Chk	0.099	22:14:10
Air Blank	0.000	22:14:50
Subject Test 1	DEF*	22:18:30
Air Blank	0.000	22:19:16
Air Blank	0.000	22:19:56
Subject Test 2	0.000	22:20:31
Air Blank	0.000	22:21:22
Wait		22:25:25
Air Blank	0.000	22:26:02
Subject Test 3	INV**	22:26:27
Air Blank	RFI***	22:26:40

Permit #: 30134

RESULTS

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004108
Test Date: 10/15/2025
Location:

Test Time: 22:27

Software Version: 9439.01.00

Standard Lot #: 302403183349
Standard Value: 0.100
Expiration Date: 01/24/2027
Bottle #: 13
Last Changed By: BLAKLEY, SCOTT
Permit #: 30134

Operator: KAUFMAN, JASON
Permit #: 30402
Agency: YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	22:27:36
Calibration Chk	0.097	22:27:58
Air Blank	0.000	22:28:38

No RFI Detected