

25-051856

Copy

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST S. Blakley AGENCY YCSO
DATE 10-29-2025 TIME 09:25:25
INTOXILYZER SERIAL # 90-0041108

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0.101 AC
- 6. Air Blanks Completed
- 7. Timer Reset


Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

Beyond 30 day No Faults Noted
Taken out of Service for PM

SIGNATURE



25 051856

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004108
Software Version: 9439.01.00
Analytical Report Number: 41081029250169

QAS: KAUFMAN, JASON
QAS Permit #: 30402
Agency: YAVAPAI CSO
Last 31-Day Check: 10/15/2025
Last Annual Maintenance: 10/30/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 10/29/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 09:10
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: BLAKLEY, SCOTT
Agency: YAVAPAI CSO

Permit #: 30134

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302403183349
Expiration Date: 01/24/2027
Bottle #: 13
Last Changed By: BLAKLEY, SCOTT

Test	g/210L	Time
Air Blank	0.000	09:25:25
Diagnostic Test	PASS	09:26:01
Air Blank	0.000	09:26:38
Calibration Chk	0.101	09:27:00
Air Blank	0.000	09:27:39
Subject Test 1	DEF*	09:32:22
Air Blank	0.000	09:33:09
Air Blank	0.000	09:33:49
Subject Test 2	0.000	09:34:15
Air Blank	0.000	09:34:58
Wait		09:39:03
Air Blank	0.000	09:39:40
Subject Test 3	INV**	09:40:28
Air Blank	RFI***	09:40:51

Permit #: 30134

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

BEYOND 30 DAY NO FAULTS NOTED

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be ±10%.*