

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST McKenna Smith AGENCY Yavapai cso
DATE 06/18/2026 TIME 1308
INTOXILYZER SERIAL # 90-004108

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0.099 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES x NO

COMMENTS:

Previous standard incorrect. Bottle lot number and bottle number changed.

SIGNATURE  12712

26-030415

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-004108

Test Date: 06/18/2026

Test Time: 13:26

Location:

Software Version: 9439.01.00

Standard Lot #: 302-403252412

Standard Value: 0.100

Expiration Date: 04/09/2027

Bottle #: 23

Last Changed By: SMITH, MCKENNA

Permit #: 30459

Operator: SMITH, MCKENNA

Permit #: 30459

Agency: YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	13:26:25
Calibration Chk	0.101	13:26:48
Air Blank	0.000	13:27:28

No RFI Detected

Successfully Completed Test Sequence

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004108
Software Version: 9439.01.00
Analytical Report Number: 41080618260226

QAS: SMITH, MCKENNA
QAS Permit #: 30459
Agency: YAVAPAI CSO
Last 31-Day Check: 05/29/2026
Last Annual Maintenance: 11/10/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 06/18/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 12:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: SMITH, MCKENNA
Agency: YAVAPAI CSO

Permit #: 30459

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302
Expiration Date: 04/09/2027
Bottle #: 403252412
Last Changed By: MCLAIN, EVAN

Permit #: 30404

Test	g/210L	Time
Air Blank	0.000	13:08:53
Diagnostic Test	PASS	13:09:28
Air Blank	0.000	13:10:05
Calibration Chk	0.099	13:10:27
Air Blank	0.000	13:11:06
Subject Test 1	DEF*	13:14:18
Air Blank	0.000	13:16:05
Air Blank	0.000	13:16:47
Subject Test 2	0.000	13:17:10
Air Blank	0.000	13:18:03
Wait		13:22:07
Air Blank	0.000	13:22:44
Subject Test 3	INV**	13:23:16
Air Blank	RFI***	13:23:37

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTSEXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.

Consecutive subject test results must not differ by more than 0.020 g/210L.

Standard check results must be $\pm 10\%$.