

PROPERTY INVOICE AND RECEIPT

NOTE: FILL IN EVERY SPACE UNLESS THE INFORMATION IS NOT AVAILABLE PRINT LEGIBLY IN INK.		TYPE OF PROPERTY (Circle One Only)		DATE AND TIME PROPERTY SEIZED	
		EVIDENCE	SAFEKEEPING	FOUND	7/3/13
NAME OF VICTIM OR OWNER (LAST, FIRST, MIDDLE)			ADDRESS		TELEPHONE
CITY OF COTTONWOOD					13-008434
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)			ADDRESS		TELEPHONE
					D.O.B.
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)			ADDRESS		TELEPHONE
					D.O.B.
NAME OF FINDER OR PERSON REPORTING FOUND PROPERTY			ADDRESS		TELEPHONE
					CITATION #
CHARGES				FINDER WILL CLAIM PROPERTY	
				YES	NO
				YES	NO

NOTIFICATION FOR RETURN OF PERSONAL PROPERTY

You are responsible for claiming your personal property within 90 days from the above date. The property will be disposed of by the Cottonwood Police Department, if no claim is made prior to that date.

ADVANCE AUTHORIZATION IS REQUIRED IF YOU WISH TO DESIGNATE ANOTHER PERSON TO RECEIVE YOUR PROPERTY.

SIGNATURE

BRIEF SYNOPSIS OF OCCURRENCE: (Also use this space for additional suspects or victims)			
90 DAY SQAP			
ITEMIZE AND DESCRIBE PROPERTY			
LIST ONLY ONE ITEM PER LINE. NUMBER EACH ITEM ON THE FORM. - Where possible, place each item in plastic bag, and attach property tag to outside of bag.			
ITEM	DESCRIPTION	SERIAL #	PROPERTY #
1 st	SQAP PAPERWORK	80-004765	89832
Placed in Property Locker #		Date	Time
6		7/3/13	1702
OFFICER:		SERIAL:	PAGE
OFC [Signature]		592	1 OF 1
DISPOSITION OF PROPERTY:			
Property Items Were:			
1. Returned to Owner		3. Sold at Auction	
2. Destroyed		4. T.O.T. City	
5. Left at Court		6. Unable to Locate	
Received By	Date	Time	Remarks
[Signature]	7/5/13	0940	
Property Location			
[Signature]			

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 15:54:26

Standard Lot#: 26212100R1
Last Changed By: R. SCARIM #592

DOS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Depuration Period? Yes

Test	g/210L	Time
RfR Blank	0.000	15:55:57
Diagnostic Test	Pass	15:56:33
RfR Blank	0.000	15:57:00
0.100 Cal Check	0.098	15:57:21
RfR Blank	0.000	15:57:51
Subject Test	0.000	15:58:07
RfR Blank	0.000	15:58:35
Wait	ABT*	
RfR Blank	0.000	15:59:09

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 15:54:26

Standard Lot#: 26212100R1
Last Changed By: R. SCARIM #592

DOS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Depuration Period? Yes

Test	g/210L	Time
RfR Blank	0.000	15:55:57
Diagnostic Test	Pass	15:56:33
RfR Blank	0.000	15:57:00
0.100 Cal Check	0.098	15:57:21
RfR Blank	0.000	15:57:51
Subject Test	0.000	15:58:07
RfR Blank	0.000	15:58:35
Wait	ABT*	
RfR Blank	0.000	15:59:09

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 15:54:26

Standard Lot#: 26212100R1
Last Changed By: R. SCARIM #592

DOS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Depuration Period? Yes

Test	g/210L	Time
RfR Blank	0.000	15:55:57
Diagnostic Test	Pass	15:56:33
RfR Blank	0.000	15:57:00
0.100 Cal Check	0.098	15:57:21
RfR Blank	0.000	15:57:51
Subject Test	0.000	15:58:07
RfR Blank	0.000	15:58:35
Wait	ABT*	
RfR Blank	0.000	15:59:09

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 16:00:07

Standard Lot#: 26212100R1
Last Changed By: R. SCARIM #592

DOS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Depuration Period? Yes

Test	g/210L	Time
RfR Blank	IPS*	16:00:49
RfR Blank	0.000	16:01:18

*Improper Sample

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 16:00:07

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Depriuation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	16:00:49
Air Blank	0.000	16:01:18

*Improper Sample

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 16:00:07

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Depriuation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	16:00:49
Air Blank	0.000	16:01:18

*Improper Sample

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 16:02:35

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Depriuation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	16:03:18
Air Blank	0.000	16:03:47

*RFI Detect

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 16:02:35

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Depriuation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	16:03:18
Air Blank	0.000	16:03:47

*RFI Detect

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 16:02:35

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test g/210L Time

Air Blank RPT* 16:03:18
Air Blank 0.000 16:03:47

*RPT Detect

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 16:04:55

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test g/210L Time

Air Blank 0.000 16:05:47
Diagnostic Test Pass 16:06:23

Air Blank 0.000 16:06:51
0.100 Cal Check 0.098 16:07:12

Air Blank 0.000 16:07:41
Subject Test 0.028* 16:10:49

Air Blank 0.000 16:11:21
Air Blank 0.000 16:11:48

Subject Test XXX** 16:12:14
Air Blank 0.000 16:12:49

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 16:04:55

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test g/210L Time

Air Blank 0.000 16:05:47
Diagnostic Test Pass 16:06:23

Air Blank 0.000 16:06:51
0.100 Cal Check 0.098 16:07:12

Air Blank 0.000 16:07:41
Subject Test 0.028* 16:10:49

Air Blank 0.000 16:11:21
Air Blank 0.000 16:11:48

Subject Test XXX** 16:12:14
Air Blank 0.000 16:12:49

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 16:04:55

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test g/210L Time

Air Blank 0.000 16:05:47
Diagnostic Test Pass 16:06:23

Air Blank 0.000 16:06:51
0.100 Cal Check 0.098 16:07:12

Air Blank 0.000 16:07:41
Subject Test 0.028* 16:10:49

Air Blank 0.000 16:11:21
Air Blank 0.000 16:11:48

Subject Test XXX** 16:12:14
Air Blank 0.000 16:12:49

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

13-008434

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
07/03/2013 16:14:36

Standard Lot#: 2621210091
Last Changed By: R. SCARIM #592

Ops: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Test	g/210L	Time
RIP Blank	0.000	16:15:07
0.100 Cal Check	0.097	16:15:28
RIP Blank	0.000	16:15:57

Preventative Maintenance
Performed --- Timer Reset

13-008434

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCARIM, ROGER AGENCY COTTONWOOD PD
DATE 7/3/13 TIME 1552M 1614
INTOXILYZER SERIAL # 80-004765 LOCATION COTTONWOOD PD

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.097 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE OK [Signature] 592

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD
DATE 7/3/13 TIME 1553
INTOXILYZER SERIAL # 90-004765 LOCATION COTTONWOOD PD

(M) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (M) 1. Clock time check.
(M) 2. Date check.

OPERATIONAL TESTS

- (M) 1. Alcohol-free subject test result 0. 000 AC.
(M) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(M) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed. 0.100
(M) 4. Standard Calibration Check standard 0.097 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES M NO _____

COMMENTS

SIGNATURE off [Signature] 592