

PROPERTY INVOICE AND RECEIPT

NOTE: FILL IN EVERY SPACE UNLESS THE INFORMATION IS NOT AVAILABLE PRINT LEGIBLY IN INK.		TYPE OF PROPERTY (Circle One Only)		DATE AND TIME PROPERTY SEIZED	
		EVIDENCE	SAFEKEEPING	FOUND	1049
NAME OF VICTIM OR OWNER (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE	D.R. NUMBER
CITY OF COTTONWOOD					13-012623
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE	D.O.B.
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE	D.O.B.
NAME OF FINDER OR PERSON REPORTING FOUND PROPERTY		ADDRESS		TELEPHONE	CITATION #
CHARGES				FINDER WILL CLAIM PROPERTY	
				YES NO	YES NO

NOTIFICATION FOR RETURN OF PERSONAL PROPERTY

You are responsible for claiming your personal property within 90 days from the above date. The property will be disposed of by the Cottonwood Police Department, if no claim is made prior to that date.

ADVANCE AUTHORIZATION IS REQUIRED IF YOU WISH TO DESIGNATE ANOTHER PERSON TO RECEIVE YOUR PROPERTY.

SIGNATURE _____

BRIEF SYNOPSIS OF OCCURRENCE: (Also use this space for additional suspects or victims)			
90 DAY SOAP			
ITEMIZE AND DESCRIBE PROPERTY			
LIST ONLY ONE ITEM PER LINE. NUMBER EACH ITEM ON THE FORM. - Where possible, place each item in plastic bag, and attach property tag to outside of bag.			
ITEM	DESCRIPTION	SERIAL #	PROPERTY #
1 st	SOAP PAPERWORK	80-004765	909D3
Placed In Property Locker # 3		Date 9/26/13	Time 1115
OFFICER: <i>[Signature]</i>		SERIAL: 592	PAGE 1 OF 1
DISPOSITION OF PROPERTY:			
Property Items Were:			
1. Returned to Owner		3. Sold at Auction	
2. Destroyed		5. Left at Court	
4. T.O.T. City		6. Unable to Locate	
Received By <i>[Signature]</i>	Date 9/27/13	Time 0805	Remarks
Property Location <i>[Signature]</i>			

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SCANNIM, ROGEN G. AGENCY COTTONWOOD PD

DATE 9/26/13 TIME 1042

INTOXILYZER SERIAL # 80-004765 LOCATION COTTONWOOD PD

(M) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(M) 1. Clock time check.

(M) 2. Date check.

OPERATIONAL TESTS

(M) 1. Alcohol-free subject test result 0. 000 AC.

(M) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed

(M) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.

(M) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES M NO

COMMENTS

SIGNATURE off [Signature] 592

13-012623

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD
DATE 9/26/13 TIME 1105
INTOXILYZER SERIAL # 80-004765 LOCATION COTTONWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (M) 4. Air blank completed.
- (M) 5. Calibration check completed. Test results 0. 097 AC.
- (M) 6. Air blank completed.
- (M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE OR M. A. 592

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:44:35

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:45:46
Diagnostic Test	Pass	10:46:23
Air Blank	0.000	10:46:50
0.100 Cal Check	0.097	10:47:11
Air Blank	0.000	10:47:41
Subject Test	0.000	10:47:57
Air Blank	0.000	10:48:24
Wait	ABT*	
Air Blank	0.000	10:49:01

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:44:35

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:45:46
Diagnostic Test	Pass	10:46:23
Air Blank	0.000	10:46:50
0.100 Cal Check	0.097	10:47:11
Air Blank	0.000	10:47:41
Subject Test	0.000	10:47:57
Air Blank	0.000	10:48:24
Wait	ABT*	
Air Blank	0.000	10:49:01

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:44:35

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:45:46
Diagnostic Test	Pass	10:46:23
Air Blank	0.000	10:46:50
0.100 Cal Check	0.097	10:47:11
Air Blank	0.000	10:47:41
Subject Test	0.000	10:47:57
Air Blank	0.000	10:48:24
Wait	ABT*	
Air Blank	0.000	10:49:01

*Sequence Aborted

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:50:13

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	10:51:12
Air Blank	0.000	10:51:41

*Improper Sample

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:50:13

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	10:51:12
Air Blank	0.000	10:51:41

*Improper Sample

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:50:13

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	10:51:12
Air Blank	0.000	10:51:41

*Improper Sample

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:52:54

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:53:51
Air Blank	0.000	10:54:19

*RFI Detect

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:52:54

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:53:51
Air Blank	0.000	10:54:19

*RFI Detect

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:52:54

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	10:53:51
Air Blank	0.000	10:54:19

*RFI Detect

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:55:12

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:56:17
Diagnostic Test	Pass	10:56:54
Air Blank	0.000	10:57:21
0.100 Cal Check	0.097	10:57:42
Air Blank	0.000	10:58:12
Subject Test	0.063*	11:01:19
Air Blank	0.000	11:01:51
Air Blank	0.000	11:02:19
Subject Test	XXX**	11:02:44
Air Blank	0.000	11:03:17

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:55:12

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:56:17
Diagnostic Test	Pass	10:56:54
Air Blank	0.000	10:57:21
0.100 Cal Check	0.097	10:57:42
Air Blank	0.000	10:58:12
Subject Test	0.063*	11:01:19
Air Blank	0.000	11:01:51
Air Blank	0.000	11:02:19
Subject Test	XXX**	11:02:44
Air Blank	0.000	11:03:17

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013 10:55:12

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

OAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SOAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	10:56:17
Diagnostic Test	Pass	10:56:54
Air Blank	0.000	10:57:21
0.100 Cal Check	0.097	10:57:42
Air Blank	0.000	10:58:12
Subject Test	0.063*	11:01:19
Air Blank	0.000	11:01:51
Air Blank	0.000	11:02:19
Subject Test	XXX**	11:02:44
Air Blank	0.000	11:03:17

*Deficient Sample
**Invalid Sample
Begin new deprivation period

Not a Successfully
Completed
Test Sequence

13-012623

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
09/26/2013

11:05:25

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	11:06:56
0.100 Cal Check	0.097	11:07:17
Air Blank	0.000	11:07:46

Preventative Maintenance
Performed -- Timer Reset

13-012623