

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCANN, ROGER G. AGENCY COTTONWOOD PD
DATE 1/9/14 TIME 0843
INTOXILYZER SERIAL # 80-004765 LOCATION COTTONWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(M) 4. Air blank completed.
(M) 5. Calibration check completed. Test results 0.103 AC.
(M) 6. Air blank completed.
(M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE OR [Signature] 592

DPS Form Exh G-2 (Rev 05-01)

14-000353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
01/09/2014

08:46:41

Standard Lot#: 26212100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	08:47:13
0.100 Cal Check	0.103	08:47:34
Air Blank	0.000	08:48:03

Preventative Maintenance
Performed -- Timer Reset

14-000 353