

PROPERTY INVOICE AND RECEIPT

NOTE: FILL IN EVERY SPACE UNLESS THE INFORMATION IS NOT AVAILABLE. PRINT LEGIBLY IN INK.		TYPE OF PROPERTY (Circle One Only)	DATE AND TIME PROPERTY SEIZED	
		<u>EVIDENCE</u>	SAFEKEEPING	FOUND
		6/28/14		1812
NAME OF VICTIM OR OWNER (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE
				D.R. NUMBER
				14-007835
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE
				D.O.B.
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE
				D.O.B.
NAME OF FINDER OR PERSON REPORTING FOUND PROPERTY		ADDRESS		TELEPHONE
				CITATION #
CHARGES			FINDER WILL CLAIM PROPERTY	
			YES	NO
			YES	NO

NOTIFICATION FOR RETURN OF PERSONAL PROPERTY

You are responsible for claiming your personal property within 90 days from the above date. The property will be disposed of by the Cottonwood Police Department, if no claim is made prior to that date.

ADVANCE AUTHORIZATION IS REQUIRED IF YOU WISH TO DESIGNATE ANOTHER PERSON TO RECEIVE YOUR PROPERTY.

SIGNATURE _____

BRIEF SYNOPSIS OF OCCURRENCE: (Also use this space for additional suspects or victims)			
31 DAY CALIBRATION CHECK			
ITEMIZE AND DESCRIBE PROPERTY			
LIST ONLY ONE ITEM PER LINE. NUMBER EACH ITEM ON THE FORM. - Where possible, place each item in plastic bag, and attach property tag to outside of bag.			
ITEM	DESCRIPTION	SERIAL #	PROPERTY #
1	CALIBRATION CHECK PAPERWORK	80-004765	94758
Placed in Property Locker #		Date	Time
		6/28/14	1812
OFFICER:	SERIAL:	PAGE	OF
officer [Signature]	592	1	1
DISPOSITION OF PROPERTY:			
Property Items Were:			
1. Returned to Owner		3. Sold at Auction	
2. Destroyed		5. Left at Court	
4. T.O.T. City		6. Unable to Locate	
Received By	Date	Time	Remarks
[Signature]	6/30/14	0900	
Property Location			
DWT			

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD
DATE 6/28/14 TIME 1759
INTOXILYZER SERIAL # 80-004765 LOCATION COTTONWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (M) 4. Air blank completed.
- (M) 5. Calibration check completed. Test results 0.101 AC.
- (M) 6. Air blank completed.
- (M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE OR. [Signature] 592

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-004765

Core Version: 8105.48

06/28/2014

18:03:37

Standard Lot#: 26212100A1

Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592

COTTONWOOD PD

Operator: R. SCARIM #592

COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	18:04:09
0.100 Cal Check	0.101	18:04:29
Air Blank	0.000	18:04:59

Preventative Maintenance
Performed -- Timer Reset

14-067835