

# PROPERTY INVOICE AND RECEIPT

NOTE: FILL IN EVERY SPACE UNLESS THE INFORMATION IS NOT AVAILABLE. PRINT LEGIBLY IN INK.		TYPE OF PROPERTY (Circle One Only)	DATE AND TIME PROPERTY SEIZED	
		<input checked="" type="radio"/> EVIDENCE <input type="radio"/> SAFEKEEPING <input type="radio"/> FOUND	10/26/14	0445
NAME OF VICTIM OR OWNER (LAST, FIRST, MIDDLE)	ADDRESS	TELEPHONE	D.R. NUMBER	
			14-013085	
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)	ADDRESS	TELEPHONE	D.O.B.	
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)	ADDRESS	TELEPHONE	D.O.B.	
NAME OF FINDER OR PERSON REPORTING FOUND PROPERTY	ADDRESS	TELEPHONE	CITATION #	
CHARGES:		FINDER WILL CLAIM PROPERTY		LAB TEST
		YES    NO	YES    NO	

### NOTIFICATION FOR RETURN OF PERSONAL PROPERTY

You are responsible for claiming your personal property within 90 days from the above date. The property will be disposed of by the Cottonwood Police Department, if no claim is made prior to that date.

ADVANCE AUTHORIZATION IS REQUIRED IF YOU WISH TO DESIGNATE ANOTHER PERSON TO RECEIVE YOUR PROPERTY.

\_\_\_\_\_  
SIGNATURE

BRIEF SYNOPSIS OF OCCURRENCE: (Also use this space for additional suspects or victims)			
90 DAY SOAP			
ITEMIZE AND DESCRIBE PROPERTY			
LIST ONLY ONE ITEM PER LINE. NUMBER EACH ITEM ON THE FORM. - Where possible, place each item in plastic bag, and attach property tag to outside of bag.			
ITEM	DESCRIPTION	SERIAL #	PROPERTY #
17	SOAP PAPER WORK	40-004765	910492
Placed in Property Locker # 12		Date 10/20/14	Time 0433
OFFICER: <i>[Signature]</i>		SERIAL 592	PAGE 1 OF 1
DISPOSITION OF PROPERTY:			
Property Items Were:			
1. Returned to Owner		3. Sold at Auction	
2. Destroyed		4. T.O.T. City	
5. Left at Court		6. Unable to Locate	
Received By	Date	Time	Remarks
<i>[Signature]</i>	10/20/14	0900	
Property Location			
DWR			

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Scaris, Roger G. AGENCY COTTONWOOD PD  
DATE 10/20/14 TIME 0418  
INTOXILYZER SERIAL # 80-004765 LOCATION COTTONWOOD PD

- (1) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (1) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(1) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(1) 4. Air blank completed.  
(1) 5. Calibration check completed. Test results 0.099 AC.  
(1) 6. Air blank completed.  
(1) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE ofc [Signature] 592

DPS Form Exh G-2 (Rev 05-01)

14-013085

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SCARIM, ROBERT G. AGENCY COTTONWOOD PD  
DATE 10/20/14 TIME 0354  
INTOXILYZER SERIAL # 80-004765 LOCATION COTTONWOOD PD

(M) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (M) 1. Clock time check.  
(M) 2. Date check.

OPERATIONAL TESTS

- (M) 1. Alcohol-free subject test result 0. 000 AC.  
(M) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(M) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(M) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 0.099 AC.

Instrument is operating properly and accurately. YES M NO     

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE OR M/M 592

14-013085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:01:10

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	04:02:17
Diagnostic Test	Pass	04:02:51
Air Blank	0.000	04:03:18
0.100 Cal Check	0.099	04:03:39
Air Blank	0.000	04:04:08
Subject Test	0.000	04:04:27
Air Blank	0.000	04:04:55
Wait	ABT*	
Air Blank	0.000	04:05:29

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

14-013085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:01:10

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	04:02:17
Diagnostic Test	Pass	04:02:51
Air Blank	0.000	04:03:18
0.100 Cal Check	0.099	04:03:39
Air Blank	0.000	04:04:08
Subject Test	0.000	04:04:27
Air Blank	0.000	04:04:55
Wait	ABT*	
Air Blank	0.000	04:05:29

\*Sequence Aborted

Not a Successfully Completed Test Sequence

14-013085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:01:10

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	04:02:17
Diagnostic Test	Pass	04:02:51
Air Blank	0.000	04:03:18
0.100 Cal Check	0.099	04:03:39
Air Blank	0.000	04:04:08
Subject Test	0.000	04:04:27
Air Blank	0.000	04:04:55
Wait	ABT*	
Air Blank	0.000	04:05:29

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

14-013085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:06:14

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	04:06:53
Air Blank	0.000	04:07:22

\*Improper Sample

Not a Successfully Completed Test Sequence

14.013085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:06:14

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	04:06:53
Air Blank	0.000	04:07:22

\*Improper Sample

**Not a Successfully Completed Test Sequence**

14-01308J

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:06:14

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	04:06:53
Air Blank	0.000	04:07:22

\*Improper Sample

Not a Successfully Completed Test Sequence

14-013085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:08:10

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	04:08:51
Air Blank	0.000	04:09:20

\*RFI Detect

Not a Successfully Completed Test Sequence

14-013085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:08:10

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP, TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	04:08:51
Air Blank	0.000	04:09:20

\*RFI Detect

Not a Successfully Completed Test Sequence

14-013085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:08:10

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	04:08:51
Air Blank	0.000	04:09:20

\*RFI Detect

Not a Successfully Completed Test Sequence

14-013085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:09:58

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	04:10:54
Diagnostic Test	Pass	04:11:28
Air Blank	0.000	04:11:55
0.100 Cal Check	0.099	04:12:16
Air Blank	0.000	04:12:46
Subject Test	0.018*	04:15:53
Air Blank	0.000	04:16:25
Air Blank	0.000	04:16:52
Subject Test	XXX**	04:17:17
Air Blank	0.000	04:17:50

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

14-013085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:09:58

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	04:10:54
Diagnostic Test	Pass	04:11:28
Air Blank	0.000	04:11:55
0.100 Cal Check	0.099	04:12:16
Air Blank	0.000	04:12:46
Subject Test	0.018*	04:15:53
Air Blank	0.000	04:16:25
Air Blank	0.000	04:16:52
Subject Test	XXX**	04:17:17
Air Blank	0.000	04:17:50

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

14-013 085

INTOXILYZER 8000  
Location: COTTONWOOD PD  
Serial Number: 80-004765  
Core Version: 8105.48  
10/20/2014

04:09:58

---

Standard Lot#: 26212100A1  
Last Changed By: R. SCARIM #592

---

QAS: R. SCARIM #592  
COTTONWOOD PD

---

Operator: R. SCARIM #592  
COTTONWOOD PD

---

Subject: SQAP,TEST  
DOB: 11/11/1911  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	04:10:54
Diagnostic Test	Pass	04:11:28
Air Blank	0.000	04:11:55
0.100 Cal Check	0.099	04:12:16
Air Blank	0.000	04:12:46
Subject Test	0.018*	04:15:53
Air Blank	0.000	04:16:25
Air Blank	0.000	04:16:52
Subject Test	XXX**	04:17:17
Air Blank	0.000	04:17:50

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

14-0130 85

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-004765

Core Version: 8105.48

10/20/2014

04:19:34

-----  
Standard Lot#: 26212100A1

Last Changed By: R. SCARIM #592  
-----

QAS: R. SCARIM #592

COTTONWOOD PD  
-----

Operator: R. SCARIM #592

COTTONWOOD PD  
-----

Test	g/210L	Time
Air Blank	0.000	04:20:05
0.100 Cal Check	0.099	04:20:25
Air Blank	0.000	04:20:55

Preventative Maintenance  
Performed -- Timer Reset

14-013085