

PROPERTY INVOICE AND RECEIPT

NOTE: FILL IN EVERY SPACE UNLESS THE INFORMATION IS NOT AVAILABLE. PRINT LEGIBLY IN INK.		TYPE OF PROPERTY (Circle One Only)	DATE AND TIME PROPERTY SEIZED	
		EVIDENCE <u>SAFEKEEPING</u> FOUND	0915 4/15/15	
NAME OF VICTIM OR OWNER (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE
N/A				15-4441
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE
				D.O.B.
NAME OF PERSON ARRESTED, OR SUSPECT (LAST, FIRST, MIDDLE)		ADDRESS		TELEPHONE
				D.O.B.
NAME OF FINDER OR PERSON REPORTING FOUND PROPERTY		ADDRESS		TELEPHONE
JIM MAKUCH, OWNER #575		199 S. 6 TH ST. COTTONWOOD AZ 86316		CITATION #
CHARGES			FINDER WILL CLAIM PROPERTY	LAB TEST
SAFEKEEPING, INTOXLYZER PAPERWORK			YES <u>NO</u>	YES <u>NO</u>

NOTIFICATION FOR RETURN OF PERSONAL PROPERTY

You are responsible for claiming your personal property within 90 days from the above date. The property will be disposed of by the Cottonwood Police Department, if no claim is made prior to that date.

ADVANCE AUTHORIZATION IS REQUIRED IF YOU WISH TO DESIGNATE ANOTHER PERSON TO RECEIVE YOUR PROPERTY.

SIGNATURE

BRIEF SYNOPSIS OF OCCURRENCE: (Also use this space for additional suspects or victims)

90 DAY SQAP TEST RESULTS.

ITEMIZE AND DESCRIBE PROPERTY

LIST ONLY ONE ITEM PER LINE. NUMBER EACH ITEM ON THE FORM. - Where possible, place each item in plastic bag, and attach property tag to outside of bag.

ITEM	DESCRIPTION	SERIAL #	PROPERTY #
1	INSTRUMENT TEST PAPERWORK #80-004765		99164

Placed in Property Locker #	011	Date	4/15/15	Time	0957
OFFICER:	OWNER J. MAKUCH	SERIAL:	595	PAGE	1 OF 1

DISPOSITION OF PROPERTY:			
Property Items Were:			
1. Returned to Owner	2. Destroyed	3. Sold at Auction	5. Left at Court
4. T.O.T. City	6. Unable to Locate		
Received By	Date	Time	Remarks
[Signature]	4/15/15	0959	
Property Location			
5 DWR			

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Debrae Hughes AGENCY Camp Verde Marshal's Office
DATE 4-15-15 TIME 0845
INTOXILYZER SERIAL # 80-004765 LOCATION Cottonwood PD

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.100 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  #53

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-004765

Core Version: 8105.48

04/15/2015

08:45:40

Standard Lot#: 09113100A1

Last Changed By: R. SCARIM #592

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	08:46:19
0.100 Cal Check	0.100	08:46:40
Air Blank	0.000	08:47:10

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshal's Ofc
DATE 4-15-15 TIME 0848 0852
INTOXILYZER SERIAL # 80-004765 LOCATION Cottonwood PD

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
 1. Clock time check.
 2. Date check.

OPERATIONAL TESTS
 1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 101 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE 

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-004765

Core Version: 8105.48

04/15/2015

08:54:09

Standard Lot#: 32114100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Subject: TEST,TEST

DOB: 04/15/1980

Sex: F

Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:55:14
Diagnostic Test	Pass	08:55:49
Air Blank	0.000	08:56:17
0.100 Cal Check	0.100	08:56:38
Air Blank	0.000	08:57:07
Subject Test	0.000	08:57:24
Air Blank	0.000	08:57:52
Wait		
Air Blank	0.000	09:02:52
Subject Test	ABT*	09:03:06
Air Blank	IPS**	09:03:14

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
04/15/2015

08:54:09

Standard Lot#: 32114100A1
Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53
CAMP VERDE MO

Operator: D. HUGHES #53
CAMP VERDE MO

Subject: TEST,TEST
DOB: 04/15/1980
Sex: F Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:55:14
Diagnostic Test	Pass	08:55:49
Air Blank	0.000	08:56:17
0.100 Cal Check	0.100	08:56:38
Air Blank	0.000	08:57:07
Subject Test	0.000	08:57:24
Air Blank	0.000	08:57:52
Wait		
Air Blank	0.000	09:02:52
Subject Test	ABT*	09:03:06
Air Blank	IPS**	09:03:14

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-004765

Core Version: 8105.48

04/15/2015

08:54:09

Standard Lot#: 32114100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Subject: TEST,TEST

DOB: 04/15/1980

Sex: F

Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:55:14
Diagnostic Test	Pass	08:55:49
Air Blank	0.000	08:56:17
0.100 Cal Check	0.100	08:56:38
Air Blank	0.000	08:57:07
Subject Test	0.000	08:57:24
Air Blank	0.000	08:57:52
Wait		
Air Blank	0.000	09:02:52
Subject Test	ABT*	09:03:06
Air Blank	IPS**	09:03:14

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
04/15/2015

09:04:09

Standard Lot#: 32114100A1
Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53
CAMP VERDE MO

Operator: D. HUGHES #53
CAMP VERDE MO

Subject: TEST,TEST
DOB: 04/15/1980
Sex: F Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:05:11
Diagnostic Test	Pass	09:05:46
Air Blank	0.000	09:06:14
0.100 Cal Check	0.101	09:06:35
Air Blank	0.000	09:07:04
Subject Test	0.020*	09:10:12
Air Blank	0.000	09:10:44
Air Blank	0.000	09:11:11
Subject Test	XXX**	09:11:35
Air Blank	RFI***	09:11:43

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
04/15/2015

09:04:09

Standard Lot#: 32114100A1
Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53
CAMP VERDE MO

Operator: D. HUGHES #53
CAMP VERDE MO

Subject: TEST,TEST
DOB: 04/15/1980
Sex: F Weight: 185

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	09:05:11
Diagnostic Test	Pass	09:05:46
Air Blank	0.000	09:06:14
0.100 Cal Check	0.101	09:06:35
Air Blank	0.000	09:07:04
Subject Test	0.020*	09:10:12
Air Blank	0.000	09:10:44
Air Blank	0.000	09:11:11
Subject Test	XXX**	09:11:35
Air Blank	RFI***	09:11:43

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshal's
DATE 4-15-15 TIME 0913
INTOXILYZER SERIAL # 80-004765 LOCATION Cottonwood PD

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 101 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] #53

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-004765

Core Version: 8105.48

04/15/2015

09:12:34

Standard Lot#: 32114100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	09:13:04
0.100 Cal Check	0.101	09:13:25
Air Blank	0.000	09:13:55

Preventative Maintenance
Performed -- Timer Reset