

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCAIM, ROGER G. AGENCY COTTONWOOD PD
DATE 2/1/18 TIME 1033
INTOXILYZER SERIAL # 80-004765 LOCATION COTTONWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(M) 4. Air blank completed.
(M) 5. Calibration check completed. Test results 0.097 AC.
(M) 6. Air blank completed.
(M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE OR M/M 592

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SCANNIN, ROGER G. AGENCY COTTONWOOD PD
DATE 2/1/18 TIME 1013
INTOXILYZER SERIAL # 80-004765 LOCATION COTTONWOOD PD

(M) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS


- (M) 1. Clock time check.
(M) 2. Date check.

OPERATIONAL TESTS

- (M) 1. Alcohol-free subject test result 0. 000 AC.
(M) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(M) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(M) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES M NO

COMMENTS

SIGNATURE ^{ofc}  592

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018

10:16:30

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:17:57 |
| Diagnostic Test | Pass | 10:18:32 |
| Air Blank | 0.000 | 10:19:00 |
| 0.100 Cal Check | 0.099 | 10:19:21 |
| Air Blank | 0.000 | 10:19:50 |
| Subject Test | 0.000 | 10:20:09 |
| Air Blank | 0.000 | 10:20:37 |
| Wait | ABT* | |
| Air Blank | 0.000 | 10:21:12 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

18-001353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018

10:16:30

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:17:57 |
| Diagnostic Test | Pass | 10:18:32 |
| Air Blank | 0.000 | 10:19:00 |
| 0.100 Cal Check | 0.099 | 10:19:21 |
| Air Blank | 0.000 | 10:19:50 |
| Subject Test | 0.000 | 10:20:09 |
| Air Blank | 0.000 | 10:20:37 |
| Wait | ABT* | |
| Air Blank | 0.000 | 10:21:12 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

18-001353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018

10:16:30

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:17:57 |
| Diagnostic Test | Pass | 10:18:32 |
| Air Blank | 0.000 | 10:19:00 |
| 0.100 Cal Check | 0.099 | 10:19:21 |
| Air Blank | 0.000 | 10:19:50 |
| Subject Test | 0.000 | 10:20:09 |
| Air Blank | 0.000 | 10:20:37 |
| Wait | ABT* | |
| Air Blank | 0.000 | 10:21:12 |

*Sequence Aborted

Not a Successfully Completed Test Sequence

18-001353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018

10:21:38

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | IPS* | 10:22:27 |
| Air Blank | 0.000 | 10:22:55 |

*Improper Sample

Not a Successfully Completed Test Sequence

18-001353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018 10:21:38

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | IPS* | 10:22:27 |
| Air Blank | 0.000 | 10:22:55 |

*Improper Sample

Not a Successfully Completed Test Sequence

18-001353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018

10:21:38

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | IPS* | 10:22:27 |
| Air Blank | 0.000 | 10:22:55 |

*Improper Sample

Not a Successfully Completed Test Sequence

18-001353

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-004765

Core Version: 8105.48

02/01/2018

10:23:23

Standard Lot#: 04216100A1

Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592

COTTONWOOD PD

Operator: R. SCARIM #592

COTTONWOOD PD

Subject: SQAP,TEST

DOB: 11/11/1911

Sex: M

Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 10:24:25 |
| Air Blank | 0.000 | 10:24:54 |

*RFI Detect

Not a Successfully Completed Test Sequence

18-00(353)

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018 10:23:23

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 10:24:25 |
| Air Blank | 0.000 | 10:24:54 |

*RFI Detect

Not a Successfully Completed Test Sequence

14-001353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018

10:23:23

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------|--------|----------|
| Air Blank | RFI* | 10:24:25 |
| Air Blank | 0.000 | 10:24:54 |

*RFI Detect

Not a Successfully Completed Test Sequence

18-001353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018

10:25:22

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:26:20 |
| Diagnostic Test | Pass | 10:26:55 |
| Air Blank | 0.000 | 10:27:22 |
| 0.100 Cal Check | 0.098 | 10:27:43 |
| Air Blank | 0.000 | 10:28:12 |
| Subject Test | 0.064* | 10:31:20 |
| Air Blank | 0.000 | 10:31:53 |
| Air Blank | 0.000 | 10:32:20 |
| Subject Test | XXX** | 10:32:34 |
| Air Blank | 0.000 | 10:33:10 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

18-001353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018

10:25:22

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:26:20 |
| Diagnostic Test | Pass | 10:26:55 |
| Air Blank | 0.000 | 10:27:22 |
| 0.100 Cal Check | 0.098 | 10:27:43 |
| Air Blank | 0.000 | 10:28:12 |
| Subject Test | 0.064* | 10:31:20 |
| Air Blank | 0.000 | 10:31:53 |
| Air Blank | 0.000 | 10:32:20 |
| Subject Test | XXX** | 10:32:34 |
| Air Blank | 0.000 | 10:33:10 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

18-001353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018

10:25:22

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:26:20 |
| Diagnostic Test | Pass | 10:26:55 |
| Air Blank | 0.000 | 10:27:22 |
| 0.100 Cal Check | 0.098 | 10:27:43 |
| Air Blank | 0.000 | 10:28:12 |
| Subject Test | 0.064* | 10:31:20 |
| Air Blank | 0.000 | 10:31:53 |
| Air Blank | 0.000 | 10:32:20 |
| Subject Test | XXX** | 10:32:34 |
| Air Blank | 0.000 | 10:33:10 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

18-001353

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
02/01/2018 10:34:22

Standard Lot#: 04216100A1
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 10:34:52 |
| 0.100 Cal Check | 0.097 | 10:35:13 |
| Air Blank | 0.000 | 10:35:42 |

Preventative Maintenance
Performed -- Timer Reset

18-001953