

INTOXILYZER 8000

Location: COTTONWOOD PD

Serial Number: 80-004765

Core Version: 8105.48

03/06/2018

18:23:06

Standard Lot#: 06117100A2

Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592

COTTONWOOD PD

Operator: R. SCARIM #592

COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	18:23:38
0.100 Cal Check	0.099	18:23:59
Air Blank	0.000	18:24:28

Preventative Maintenance

Performed -- Timer Reset

18-002765

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCAIM, ROGER G. AGENCY COTTENWOOD PD
DATE 3/6/18 TIME 1920
INTOXILYZER SERIAL # 80-004765 LOCATION COTTENWOOD PD

- (1) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (1) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (1) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (1) 4. Air blank completed.
- (1) 5. Calibration check completed. Test results 0.099 AC.
- (1) 6. Air blank completed.
- (1) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE ^{of} [Signature] 592