

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Scrim, ROGER G. AGENCY COTTONWOOD PD

DATE 11/1/18 TIME 1524

INTOXILYZER SERIAL # 40-004765 LOCATION COTTONWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (M) 4. Air blank completed.
- (M) 5. Calibration check completed. Test results 0.098 AC.
- (M) 6. Air blank completed.
- (M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE OFc [Signature] 592

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD

DATE 11/1/18 TIME 1553

INTOXILYZER SERIAL # 80-00 4765 LOCATION COTTONWOOD PD

- (H) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (H) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (H) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (H) 4. Air blank completed.
- (H) 5. Calibration check completed. Test results 0.097 AC.
- (H) 6. Air blank completed.
- (H) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE ofc. [Signature]

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SCANNIN, ROGER G. AGENCY COTTONWOOD PD
DATE 11/1/18 TIME 1530
INTOXILYZER SERIAL # 80-004765 LOCATION COTTONWOOD PD

(M) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (M) 1. Clock time check.
(M) 2. Date check.

OPERATIONAL TESTS

- (M) 1. Alcohol-free subject test result 0. 000 AC.
(M) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(M) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(M) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES M NO

COMMENTS INSTRUMENT REMOVED FOR ANNUAL SERVICE
WITH DPS.

SIGNATURE OFc [Signature] 597

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018 15:28:06

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	15:28:38
0.100 Cal Check	0.098	15:28:58
Air Blank	0.000	15:29:28

Preventative Maintenance
Performed -- Timer Reset

18-014042

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:36:21

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:37:25
Diagnostic Test	Pass	15:38:00
Air Blank	0.000	15:38:28
0.100 Cal Check	0.098	15:38:49
Air Blank	0.000	15:39:18
Subject Test	0.000	15:39:35
Air Blank	0.000	15:40:02
Wait	ABT*	
Air Blank	0.000	15:40:38

*Sequence Aborted

Not a Successfully Completed Test Sequence

18-014647

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:36:21

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:37:25
Diagnostic Test	Pass	15:38:00
Air Blank	0.000	15:38:28
0.100 Cal Check	0.098	15:38:49
Air Blank	0.000	15:39:18
Subject Test	0.000	15:39:35
Air Blank	0.000	15:40:02
Wait	ABT*	
Air Blank	0.000	15:40:38

*Sequence Aborted

Not a Successfully Completed Test Sequence

18-014042

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:36:21

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:37:25
Diagnostic Test	Pass	15:38:00
Air Blank	0.000	15:38:28
0.100 Cal Check	0.098	15:38:49
Air Blank	0.000	15:39:18
Subject Test	0.000	15:39:35
Air Blank	0.000	15:40:02
Wait	ABT*	
Air Blank	0.000	15:40:38

*Sequence Aborted

Not a Successfully Completed Test Sequence

18-014042

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:41:10

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	15:41:49
Air Blank	0.000	15:42:17

*Improper Sample

Not a Successfully Completed Test Sequence

18-014042

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:41:10

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	15:41:49
Air Blank	0.000	15:42:17

*Improper Sample

Not a Successfully Completed Test Sequence

18-014042

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:41:10

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	15:41:49
Air Blank	0.000	15:42:17

*Improper Sample

Not a Successfully Completed Test Sequence

18-014047

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:42:57

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	15:43:41
Air Blank	0.000	15:44:09

*RFI Detect

Not a Successfully Completed Test Sequence

18-014 047

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:42:57

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	15:43:41
Air Blank	0.000	15:44:09

*RFI Detect

Not a Successfully Completed Test Sequence

18-G14042

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:42:57

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	15:43:41
Air Blank	0.000	15:44:09

*RFI Detect

Not a Successfully Completed Test Sequence

18-014042

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:44:51

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:45:45
Diagnostic Test	Pass	15:46:20
Air Blank	0.000	15:46:48
0.100 Cal Check	0.098	15:47:08
Air Blank	0.000	15:47:38
Subject Test	0.083*	15:50:45
Air Blank	0.000	15:51:19
Air Blank	0.000	15:51:46
Subject Test	XXX**	15:51:57
Air Blank	0.000	15:52:33

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

18-014042

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:44:51

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:45:45
Diagnostic Test	Pass	15:46:20
Air Blank	0.000	15:46:48
0.100 Cal Check	0.098	15:47:08
Air Blank	0.000	15:47:38
Subject Test	0.083*	15:50:45
Air Blank	0.000	15:51:19
Air Blank	0.000	15:51:46
Subject Test	XXX**	15:51:57
Air Blank	0.000	15:52:33

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

18-014042

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:44:51

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:45:45
Diagnostic Test	Pass	15:46:20
Air Blank	0.000	15:46:48
0.100 Cal Check	0.098	15:47:08
Air Blank	0.000	15:47:38
Subject Test	0.083*	15:50:45
Air Blank	0.000	15:51:19
Air Blank	0.000	15:51:46
Subject Test	XXX**	15:51:57
Air Blank	0.000	15:52:33

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

18-014042

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/01/2018

15:53:57

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	15:54:27
0.100 Cal Check	0.097	15:54:47
Air Blank	0.000	15:55:17

Preventative Maintenance
Performed -- Timer Reset

18-014042