

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SCARIM, ROGER G. AGENCY COTTONWOOD PD
DATE 11/29/18 TIME 1642
INTOXILYZER SERIAL # 90-004765 LOCATION COTTONWOOD PD

(M) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (M) 1. Clock time check.
(M) 2. Date check.

OPERATIONAL TESTS

- (M) 1. Alcohol-free subject test result 0.000 AC.
(M) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(M) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(M) 4. Standard Calibration Check standard 0.100 AC. Result 0.097 AC.

Instrument is operating properly and accurately. YES M NO

COMMENTS

SIGNATURE 

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SCHEIM, ROGER G. AGENCY COTTOWOOD PD
DATE 11/29/18 TIME 1742
INTOXILYZER SERIAL # 80-004765 LOCATION COTTOWOOD PD

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.10% AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(M) 4. Air blank completed.
(M) 5. Calibration check completed. Test results 0.097 AC.
(M) 6. Air blank completed.
(M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE off M J R 592

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:44:13

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:45:31
Diagnostic Test	Pass	16:46:06
Air Blank	0.000	16:46:33
0.100 Cal Check	0.097	16:46:54
Air Blank	0.000	16:47:23
Subject Test	0.000	16:47:39
Air Blank	0.000	16:48:06
Wait	ABT*	
Air Blank	0.000	16:48:48

*Sequence Aborted

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:44:13

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:45:31
Diagnostic Test	Pass	16:46:06
Air Blank	0.000	16:46:33
0.100 Cal Check	0.097	16:46:54
Air Blank	0.000	16:47:23
Subject Test	0.000	16:47:39
Air Blank	0.000	16:48:06
Wait	ABT*	
Air Blank	0.000	16:48:48

*Sequence Aborted

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:44:13

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:45:31
Diagnostic Test	Pass	16:46:06
Air Blank	0.000	16:46:33
0.100 Cal Check	0.097	16:46:54
Air Blank	0.000	16:47:23
Subject Test	0.000	16:47:39
Air Blank	0.000	16:48:06
Wait	ABT*	
Air Blank	0.000	16:48:48

*Sequence Aborted

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:49:32

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	16:50:13
Air Blank	0.000	16:50:42

*Improper Sample

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:49:32

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	16:50:13
Air Blank	0.000	16:50:42

*Improper Sample

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:49:32

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	16:50:13
Air Blank	0.000	16:50:42

*Improper Sample

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:51:07

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP, TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	16:52:01
Air Blank	0.000	16:52:29

*RFI Detect

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:51:07

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	16:52:01
Air Blank	0.000	16:52:29

*RFI Detect

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:51:07

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	16:52:01
Air Blank	0.000	16:52:29

*RFI Detect

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:53:05

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:53:56
Diagnostic Test	Pass	16:54:31
Air Blank	0.000	16:54:59
0.100 Cal Check	0.097	16:55:19
Air Blank	0.000	16:55:49
Subject Test	RGE*	16:57:34
Air Blank	0.000	16:58:12

*Range Exceeded

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:53:05

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:53:56
Diagnostic Test	Pass	16:54:31
Air Blank	0.000	16:54:59
0.100 Cal Check	0.097	16:55:19
Air Blank	0.000	16:55:49
Subject Test	RGE*	16:57:34
Air Blank	0.000	16:58:12

*Range Exceeded

Not a Successfully Completed Test Sequence

18 - 015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

16:53:05

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:53:56
Diagnostic Test	Pass	16:54:31
Air Blank	0.000	16:54:59
0.100 Cal Check	0.097	16:55:19
Air Blank	0.000	16:55:49
Subject Test	RGE*	16:57:34
Air Blank	0.000	16:58:12

*Range Exceeded

Not a Successfully Completed Test Sequence

18-015216

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

17:33:20

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:34:47
Diagnostic Test	Pass	17:35:22
Air Blank	0.000	17:35:49
0.100 Cal Check	0.098	17:36:10
Air Blank	0.000	17:36:40
Subject Test	0.194*	17:39:47
Air Blank	0.000	17:40:19
Air Blank	0.000	17:40:47
Subject Test	XXX**	17:41:06
Air Blank	0.000	17:41:40

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

17:33:20

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:34:47
Diagnostic Test	Pass	17:35:22
Air Blank	0.000	17:35:49
0.100 Cal Check	0.098	17:36:10
Air Blank	0.000	17:36:40
Subject Test	0.194*	17:39:47
Air Blank	0.000	17:40:19
Air Blank	0.000	17:40:47
Subject Test	XXX**	17:41:06
Air Blank	0.000	17:41:40

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

17:33:20

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Subject: SQAP,TEST
DOB: 11/11/1911
Sex: M Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:34:47
Diagnostic Test	Pass	17:35:22
Air Blank	0.000	17:35:49
0.100 Cal Check	0.098	17:36:10
Air Blank	0.000	17:36:40
Subject Test	0.194*	17:39:47
Air Blank	0.000	17:40:19
Air Blank	0.000	17:40:47
Subject Test	XXX**	17:41:06
Air Blank	0.000	17:41:40

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: COTTONWOOD PD
Serial Number: 80-004765
Core Version: 8105.48
11/29/2018

17:43:17

Standard Lot#: 06117100A2
Last Changed By: R. SCARIM #592

QAS: R. SCARIM #592
COTTONWOOD PD

Operator: R. SCARIM #592
COTTONWOOD PD

Test	g/210L	Time
Air Blank	0.000	17:43:46
0.100 Cal Check	0.097	17:44:07
Air Blank	0.000	17:44:36

Preventative Maintenance
Performed -- Timer Reset