

EXHIBIT I-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST JAKE CLEVENGER AGENCY CAMP VERDE MARSHAL'S OFFICE  
DATE 3-2-26 TIME 2128  
INTOXILYZER SERIAL # 90-005305

1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

1. Clock time check  
 2. Date check

OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
5. Standard Calibration Check:  
0.104 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES  NO

COMMENTS:

SIGNATURE

 #200

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-005305  
Software Version: 9439.01.00  
Analytical Report Number: 53050302260108

QAS: CLEVINGER, JACOB  
QAS Permit #: 30330  
Agency: CAMP VERDE MO  
Last 31-Day Check: 01/27/2026  
Last Annual Maintenance: 06/19/2025

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 03/02/2026  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 21:00  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: CLEVINGER, JACOB  
Agency: CAMP VERDE MO

Permit #: 30330

## STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 302403337473  
Expiration Date: 06/13/2027  
Bottle #: 11  
Last Changed By: CLEVINGER, JACOB

Permit #: 30330

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	21:29:36
Diagnostic Test	PASS	21:30:11
Air Blank	0.000	21:30:48
Calibration Chk	0.104	21:31:10
Air Blank	0.000	21:31:50
Subject Test 1	DEF*	21:35:21
Air Blank	0.000	21:36:06
Air Blank	0.000	21:36:46
Subject Test 2	0.000	21:37:12
Air Blank	0.000	21:37:55
Wait		21:41:59
Air Blank	0.000	21:42:36
Subject Test 3	INV**	21:43:07
Air Blank	RFI***	21:43:20

## RESULTS

## EXCEPTION MESSAGES

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be ±10%.*