

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST JAKE CLEVENGER AGENCY CAMP VERDE MARSHAL'S
DATE 6-24-26 TIME 0400
INTOXILYZER SERIAL # 90-005305

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0. 101 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

BACK IN SERVICE AFTER ANNUAL MAINT.

SIGNATURE

#200

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-005305
Software Version: 9439.01.00
Analytical Report Number: 53050624260118

QAS: CLEVINGER, JACOB
QAS Permit #: 30330
Agency: CAMP VERDE MO
Last 31-Day Check: 04/02/2026
Last Annual Maintenance: 05/18/2026

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 06/24/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 03:45
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: CLEVINGER, JACOB
Agency: CAMP VERDE MO

Permit #: 30330

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302403337473
Expiration Date: 06/13/2027
Bottle #: 11
Last Changed By: CLEVINGER, JACOB

Permit #: 30330

Test	g/210L	Time
Air Blank	0.000	04:02:55
Diagnostic Test	PASS	04:03:30
Air Blank	0.000	04:04:07
Calibration Chk	0.101	04:04:29
Air Blank	0.000	04:05:11
Subject Test 1	DEF*	04:08:45
Air Blank	0.000	04:09:31
Air Blank	0.000	04:10:11
Subject Test 2	0.000	04:10:37
Air Blank	0.000	04:11:18
Wait		04:15:22
Air Blank	0.000	04:15:59
Subject Test 3	INV**	04:16:38
Air Blank	RFI***	04:16:53

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*