

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST P. Reed AGENCY PPD
DATE 1/11/2026 TIME 1854
INTOXILYZER SERIAL # 90-005776

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.10 AC

DIAGNOSTIC TESTS

1. Clock time check
 2. Date check


OPERATIONAL TESTS

1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
5. Standard Calibration Check:
0. 098 AC
6. Air Blanks Completed
7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE 

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-005776
Software Version: 9439.01.00
Analytical Report Number: 57760101260195

QAS: REED, PERRY
QAS Permit #: 30248
Agency: PRESCOTT PD
Last 31-Day Check: 12/24/2025
Last Annual Maintenance: 02/24/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 01/01/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 16:39
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: REED, PERRY
Agency: PRESCOTT PD

Permit #: 30248

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 15924100A1
Expiration Date: 07/05/2026
Bottle #: 5
Last Changed By: NAMANNY, NATHANIEL

Permit #: 30250

Test	g/210L	Time
Air Blank	0.000	16:54:50
Diagnostic Test	PASS	16:55:26
Air Blank	0.000	16:56:02
Calibration Chk	0.098	16:56:24
Air Blank	0.000	16:57:04
Subject Test 1	DEF*	17:00:16
Air Blank	0.000	17:00:57
Air Blank	0.000	17:01:38
Subject Test 2	0.000	17:01:55
Air Blank	0.000	17:02:35
Wait		17:06:38
Air Blank	0.000	17:07:15
Subject Test 3	INV**	17:09:06
Air Blank	RFI***	17:09:19

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.